

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072225\
 Data File : BF143201.D
 Acq On : 22 Jul 2025 21:18
 Operator : RC/JU
 Sample : Q2649-21
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-6

Quant Time: Jul 23 04:53:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 17 15:14:05 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.969	152	107235	20.000	ng	0.00
21) Naphthalene-d8	8.245	136	362389	20.000	ng	0.00
39) Acenaphthene-d10	10.004	164	152041	20.000	ng	0.00
64) Phenanthrene-d10	11.486	188	230194	20.000	ng	0.00
76) Chrysene-d12	14.127	240	261475	20.000	ng	0.00
86) Perylene-d12	15.639	264	188557	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.604	112	464041	68.478	ng	0.02
7) Phenol-d6	6.592	99	611120	71.648	ng	0.00
23) Nitrobenzene-d5	7.522	82	328770	45.223	ng	0.00
42) 2,4,6-Tribromophenol	10.792	330	122121	77.751	ng	0.00
45) 2-Fluorobiphenyl	9.322	172	487886	43.886	ng	0.00
79) Terphenyl-d14	13.068	244	614703	34.984	ng	0.00
Target Compounds						
60) Diethylphthalate	10.410	149	70173	7.164	ng	98
74) Di-n-butylphthalate	12.045	149	26388	2.143	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072225\
Data File : BF143201.D
Acq On : 22 Jul 2025 21:18
Operator : RC/JU
Sample : Q2649-21
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-6

Quant Time: Jul 23 04:53:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 17 15:14:05 2025
Response via : Initial Calibration

