

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072315\
 Data File : BF080611.D
 Acq On : 24 Jul 2015 1:04
 Operator : TP/UM
 Sample : G3068-03MS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Quant Time: Jul 24 02:29:07 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 24 01:01:51 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.94	152	39538	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	161698	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	72575	20.00	ng	0.00
63) Phenanthrene-d10	11.48	188	124713	20.00	ng	0.00
75) Chrysene-d12	14.23	240	108351	20.00	ng	0.05
86) Perylene-d12	15.85	264	82362	20.00	ng	0.10

System Monitoring Compounds

5) 2-Fluorophenol	5.53	112	352070	153.80	ng	0.01
7) Phenol-d6	6.56	99	401032	143.31	ng	0.00
23) Nitrobenzene-d5	7.50	82	262284	91.91	ng	0.00
41) 2,4,6-Tribromophenol	10.78	330	94357	164.08	ng	0.01
44) 2-Fluorobiphenyl	9.31	172	387693	77.51	ng	0.00
78) Terphenyl-d14	13.08	244	374914	72.41	ng	0.01

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.50	88	41505	38.06	ng	# 100
3) Pyridine	3.28	79	109981	37.95	ng	93
4) n-Nitrosodimethylamine	3.18	42	74973	40.93	ng	# 97
6) Aniline	6.60	93	145206	35.50	ng	99
8) 2-Chlorophenol	6.73	128	118498	48.20	ng	97
9) Benzaldehyde	6.49	77	36355	17.57	ng	93
10) Phenol	6.57	94	141693	42.32	ng	82
11) bis(2-Chloroethyl)ether	6.67	93	103690	44.60	ng	98
12) 1,3-Dichlorobenzene	6.89	146	139176	44.65	ng	98
13) 1,4-Dichlorobenzene	6.97	146	134873	44.09	ng	99
14) 1,2-Dichlorobenzene	7.12	146	133074	46.92	ng	99
15) Benzyl Alcohol	7.08	79	122258	49.57	ng	99
16) 2,2'-oxybis(1-Chloropropan	7.23	45	224119	47.56	ng	99
17) 2-Methylphenol	7.20	107	88761	45.81	ng	93
18) Hexachloroethane	7.47	117	49202	46.50	ng	96
19) n-Nitroso-di-n-propylamine	7.36	70	91726	48.38	ng	97
20) 3+4-Methylphenols	7.34	107	128531	47.98	ng	# 82
22) Acetophenone	7.36	105	182150	46.25	ng	98
24) Nitrobenzene	7.53	77	142427	46.69	ng	98
25) Isophorone	7.77	82	237323	44.90	ng	100
26) 2-Nitrophenol	7.85	139	52864	48.50	ng	90
27) 2,4-Dimethylphenol	7.88	122	104861	43.55	ng	93
28) bis(2-Chloroethoxy)methane	7.98	93	134585	46.08	ng	99
29) 2,4-Dichlorophenol	8.09	162	94429	47.96	ng	97
30) 1,2,4-Trichlorobenzene	8.18	180	111061	43.38	ng	98
31) Naphthalene	8.26	128	355886	44.00	ng	100
32) Benzoic acid	7.94	122	36287	34.65	ng	# 79
33) 4-Chloroaniline	8.30	127	89509	28.21	ng	95
34) Hexachlorobutadiene	8.38	225	68111	45.64	ng	94
35) Caprolactam	8.64	113	24588	46.37	ng	# 74
36) 4-Chloro-3-methylphenol	8.77	107	112370	47.97	ng	95
37) 2-Methylnaphthalene	8.96	142	205253	45.31	ng	95
39) 1,2,4,5-Tetrachlorobenzene	9.12	216	107654	43.87	ng	98
40) Hexachlorocyclopentadiene	9.10	237	124524	96.35	ng	95

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072315\
 Data File : BF080611.D
 Acq On : 24 Jul 2015 1:04
 Operator : TP/UM
 Sample : G3068-03MS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Quant Time: Jul 24 02:29:07 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 24 01:01:51 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.22	196	61427	44.63	ng	99
43) 2,4,5-Trichlorophenol	9.25	196	66080	47.55	ng	93
45) 1,1'-Biphenyl	9.41	154	270852	45.85	ng	97
46) 2-Chloronaphthalene	9.44	162	212739	44.51	ng	98
47) 2-Nitroaniline	9.53	65	70423	48.32	ng	97
48) Acenaphthylene	9.86	152	324952	45.90	ng	99
49) Dimethylphthalate	9.71	163	280022	52.44	ng	100
50) 2,6-Dinitrotoluene	9.77	165	47106	47.16	ng	85
51) Acenaphthene	10.03	154	212969	50.07	ng	97
52) 3-Nitroaniline	9.94	138	40666	37.44	ng	# 93
53) 2,4-Dinitrophenol	10.04	184	32904	86.63	ng	# 78
54) Dibenzofuran	10.20	168	292334	47.56	ng	99
55) 4-Nitrophenol	10.09	139	75961	96.81	ng	# 67
56) 2,4-Dinitrotoluene	10.17	165	62904	50.28	ng	# 96
57) Fluorene	10.54	166	229693	46.54	ng	97
58) 2,3,4,6-Tetrachlorophenol	10.32	232	55317	51.16	ng	99
59) Diethylphthalate	10.41	149	226451	45.44	ng	97
60) 4-Chlorophenyl-phenylether	10.53	204	98488	44.43	ng	98
61) 4-Nitroaniline	10.54	138	50143	45.84	ng	87
62) Azobenzene	10.69	77	254382	48.73	ng	98
64) 4,6-Dinitro-2-methylphenol	10.58	198	23153	52.85	ng	81
65) n-Nitrosodiphenylamine	10.65	169	192753	47.10	ng	99
66) 4-Bromophenyl-phenylether	11.02	248	60523	47.74	ng	91
67) Hexachlorobenzene	11.09	284	69044	48.24	ng	94
68) Atrazine	11.17	200	60058	53.42	ng	98
69) Pentachlorophenol	11.28	266	67760	101.45	ng	97
70) Phenanthrene	11.50	178	340372	48.35	ng	98
71) Anthracene	11.55	178	317312	48.13	ng	99
72) Carbazole	11.71	167	293211	51.44	ng	98
73) Di-n-butylphthalate	12.04	149	326071	49.22	ng	# 98
74) Fluoranthene	12.70	202	311460	49.44	ng	94
76) Benzidine	12.82	184	128093	40.23	ng	99
77) Pyrene	12.93	202	334156	45.97	ng	99
79) Butylbenzylphthalate	13.60	149	136029	51.64	ng	98
80) Benzo(a)anthracene	14.21	228	278297	45.55	ng	99
81) 3,3'-Dichlorobenzidine	14.18	252	75760	43.39	ng	95
82) Chrysene	14.26	228	271127	44.61	ng	97
83) Bis(2-ethylhexyl)phthalate	14.23	149	185475	52.19	ng	99
84) Di-n-octyl phthalate	14.95	149	262002	52.05	ng	94
85) Indeno(1,2,3-cd)pyrene	17.36	276	281547	61.23	ng	96
87) Benzo(b)fluoranthene	15.41	252	247828	49.47	ng	# 97
88) Benzo(k)fluoranthene	15.41	252	247828	49.71	ng	# 95
89) Benzo(a)pyrene	15.79	252	222963	50.17	ng	# 91
90) Dibenzo(a,h)anthracene	17.38	278	233962	55.74	ng	98
91) Benzo(g,h,i)perylene	17.80	276	254919	59.46	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072315\
Data File : BF080611.D
Acq On : 24 Jul 2015 1:04
Operator : TP/UM
Sample : G3068-03MS
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Quant Time: Jul 24 02:29:07 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 24 01:01:51 2015
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.94 min Scan# 425
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:152 Resp: 39538
Ion Ratio Lower Upper
152 100
150 161.3 128.7 193.1
115 71.8 56.8 85.2




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 38.06 ng
RT: 2.50 min Scan# 36
Delta R.T. 0.02 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion: 88 Resp: 41505
Ion Ratio Lower Upper
88 100
58 92.3 0.0 0.0#
43 49.4 0.0 0.0#



Abundance Ion 88.00 (87.70 to 88.70): BF0
Ion 58.00 (57.70 to 58.70): BF0
Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 37.95 ng

RT: 3.28 min Scan# 104

Delta R.T. 0.02 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS





Tgt Ion: 79 Resp: 109981

Ion Ratio Lower Upper

79 100

52 94.2 79.6 119.4

51 49.1 44.5 66.7



#4

n-Nitrosodimethylamine

Concen: 40.93 ng

RT: 3.18 min Scan# 96

Delta R.T. 0.01 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04




Tgt Ion: 42 Resp: 74973


Ion Ratio Lower Upper

42 100

74 91.9 72.6 108.8

44 7.1 10.4 15.6#





#5

2-Fluorophenol

Concen: 153.80 ng

RT: 5.53 min Scan# 301

Delta R.T. 0.01 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion: 112 Resp: 352070

Ion Ratio Lower Upper

112 100

64 61.6 60.2 90.2

63 34.8 34.8 52.2#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0

500000

400000

300000



200000

100000

0

5.45 5.50 5.55 5.60

Time-->



#6

Aniline

Concen: 35.50 ng

RT: 6.60 min Scan# 395

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion: 93 Resp: 145206

Ion Ratio Lower Upper

93 100

66 42.1 33.1 49.7

65 22.4 17.2 25.8



Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

200000

150000


100000


50000

0

6.50 6.55 6.60 6.65

Time-->





#7

Phenol-d6

Concen: 143.31 ng

RT: 6.56 min Scan# 391

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion: 99 Resp: 401032

Ion Ratio Lower Upper

99 100

42 28.5 17.4 26.2#




71 43.2 29.0 43.6



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 48.20 ng

RT: 6.73 min Scan# 406

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion: 128 Resp: 118498

Ion Ratio Lower Upper

128 100

130 32.5 11.8 51.8



64 43.2 20.2 60.2




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 17.57 ng

RT: 6.49 min Scan# 385

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion: 77 Resp: 36355

Ion Ratio Lower Upper

77 100

105 100.7 72.3 112.3




106 92.9 67.5 107.5



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 42.32 ng

RT: 6.57 min Scan# 392

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion: 94 Resp: 141693

Ion Ratio Lower Upper

94 100

65 39.3 12.1 52.1



66 58.1 24.3 64.3




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0





#11

bis(2-Chloroethyl)ether

Concen: 44.60 ng

RT: 6.67 min Scan# 401

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion: 93 Resp: 103690

Ion Ratio Lower Upper

93 100

63 103.5 80.5 120.5




95 30.9 11.2 51.2



Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 63.00 (62.70 to 63.70): BF0

Ion 95.00 (94.70 to 95.70): BF0



#12

1,3-Dichlorobenzene

Concen: 44.65 ng

RT: 6.89 min Scan# 420

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion: 146 Resp: 139176

Ion Ratio Lower Upper

146 100

148 65.5 50.9 76.3



75 34.0 27.1 40.7




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 75.00 (74.70 to 75.70): BF0









#13
 1,4-Dichlorobenzene
 Concen: 44.09 ng
 RT: 6.97 min Scan# 427
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04




Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS


Tgt Ion:146 Resp: 134873
 Ion Ratio Lower Upper
 146 100
 148 66.1 52.0 78.0
 111 40.1 31.4 47.2



#14
 1,2-Dichlorobenzene
 Concen: 46.92 ng
 RT: 7.12 min Scan# 440
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:146 Resp: 133074
 Ion Ratio Lower Upper
 146 100
 148 62.4 50.3 75.5
 111 47.3 38.2 57.2





#15

Benzyl Alcohol

Concen: 49.57 ng

RT: 7.08 min Scan# 437

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion: 79 Resp: 122258

Ion Ratio Lower Upper

79 100

108 74.8 61.5 92.3




77 67.0 53.6 80.4



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 108.00 (107.70 to 108.70): BF0

Ion 77.00 (76.70 to 77.70): BF0



#16

2,2'-oxybis(1-Chloropropane)

Concen: 47.56 ng

RT: 7.23 min Scan# 450

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion: 45 Resp: 224119

Ion Ratio Lower Upper

45 100

77 11.0 0.0 30.8



79 9.0 0.0 28.6




Abundance Ion 45.00 (44.70 to 45.70): BF0

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0









#17
2-Methylphenol
Concen: 45.81 ng
RT: 7.20 min Scan# 447
Delta R.T. 0.01 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:107 Resp: 88761
Ion Ratio Lower Upper
107 100
108 111.1 86.0 129.0
77 46.0 43.4 65.2
79 44.2 41.2 61.8




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF0
Ion 79.00 (78.70 to 79.70): BF0








#18
Hexachloroethane
Concen: 46.50 ng
RT: 7.47 min Scan# 471
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:117 Resp: 49202
Ion Ratio Lower Upper
117 100
119 99.4 76.5 114.7
201 119.2 91.6 137.4



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 48.38 ng

RT: 7.36 min Scan# 461

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS

Tgt Ion: 70 Resp: 91726

Ion Ratio Lower Upper

70 100

42 71.9 56.6 85.0

101 9.9 9.7 14.5

130 20.5 19.1 28.7





Abundance

Ion 70.00 (69.70 to 70.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 101.00 (100.70 to 101.70): BF0

Ion 130.00 (129.70 to 130.70): BF0



#20

3+4-Methylphenols

Concen: 47.98 ng

RT: 7.34 min Scan# 460

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Tgt Ion: 107 Resp: 128531

Ion Ratio Lower Upper

107 100

108 87.6 75.6 115.6

77 72.2 84.7 124.7#

79 34.6 13.8 53.8


Abundance


Ion 107.00 (106.70 to 107.70): BF0

Ion 108.00 (107.70 to 108.70): BF0

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.24 min Scan# 538

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS

Tgt Ion:136 Resp: 161698


Ion Ratio Lower Upper

136 100

137 10.4 7.8 11.8

54 8.9 7.2 10.8

68 4.1 3.8 5.6






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 46.25 ng

RT: 7.36 min Scan# 461

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Tgt Ion:105 Resp: 182150


Ion Ratio Lower Upper

105 100

71 7.6 5.5 8.3

51 30.4 23.0 34.4

120 20.1 15.8 23.8






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E





#23

Nitrobenzene-d5

Concen: 91.91 ng

RT: 7.50 min Scan# 474

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion: 82 Resp: 262284

Ion Ratio Lower Upper

82 100

128 34.0 28.9 43.3




54 65.4 54.5 81.7



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): BF0

Ion 54.00 (53.70 to 54.70): BF0



#24

Nitrobenzene

Concen: 46.69 ng

RT: 7.53 min Scan# 476

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion: 77 Resp: 142427

Ion Ratio Lower Upper

77 100

123 45.2 37.1 55.7



65 14.5 11.4 17.2




Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 123.00 (122.70 to 123.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 44.90 ng

RT: 7.77 min Scan# 497

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion: 82 Resp: 237323

Ion Ratio Lower Upper




82 100

95 7.8 5.9 8.9

138 16.8 13.4 20.2



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#26

2-Nitrophenol

Concen: 48.50 ng

RT: 7.85 min Scan# 504

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion: 139 Resp: 52864

Ion Ratio Lower Upper



139 100


109 38.5 26.0 39.0

65 63.7 45.4 68.0



Abundance Ion 139.00 (138.70 to 139.70): BF0
Ion 109.00 (108.70 to 109.70): BF0
Ion 65.00 (64.70 to 65.70): BF0





#27

2,4-Dimethylphenol

Concen: 43.55 ng

RT: 7.88 min Scan# 507

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion:122 Resp: 104861

Ion Ratio Lower Upper

122 100

107 126.0 92.3 138.5




121 60.5 47.1 70.7



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 46.08 ng

RT: 7.98 min Scan# 516

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion: 93 Resp: 134585

Ion Ratio Lower Upper

93 100

95 32.2 25.9 38.9



123 11.2 10.3 15.5




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 95.00 (94.70 to 95.70): BF0

Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 47.96 ng
 RT: 8.09 min Scan# 525
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Tgt Ion	Ratio	Lower	Upper
162	100		
164	65.5	44.1	84.1
98	31.9	8.9	48.9




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF0






#30
 1,2,4-Trichlorobenzene
 Concen: 43.38 ng
 RT: 8.18 min Scan# 533
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion	Ratio	Lower	Upper
180	100		
182	96.6	79.0	118.4
145	29.0	22.0	33.0



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 44.00 ng
RT: 8.26 min Scan# 540
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:128 Resp: 355886
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.3 10.7 16.1




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 34.65 ng
RT: 7.94 min Scan# 512
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:122 Resp: 36287
Ion Ratio Lower Upper
122 100
105 119.4 112.5 152.5
77 85.7 102.5 142.5#



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 28.21 ng
RT: 8.30 min Scan# 544
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:	127	Resp:	89509
Ion	Ratio	Lower	Upper
127	100		
129	33.0	27.8	41.6
65	30.6	21.2	31.8
92	18.9	13.9	20.9




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0






#34
Hexachlorobutadiene
Concen: 45.64 ng
RT: 8.38 min Scan# 551
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:	225	Resp:	68111
Ion	Ratio	Lower	Upper
225	100		
223	59.9	50.3	75.5
227	61.5	54.5	81.7



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 46.37 ng

RT: 8.64 min Scan# 573

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion:113 Resp: 24588

Ion Ratio Lower Upper

113 100

55 220.9 259.3 299.3#




56 162.9 171.6 211.6#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#36

4-Chloro-3-methylphenol

Concen: 47.97 ng

RT: 8.77 min Scan# 585

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion:107 Resp: 112370

Ion Ratio Lower Upper

107 100

144 19.9 17.7 26.5



142 67.5 57.7 86.5




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 45.31 ng
 RT: 8.96 min Scan# 601
 Delta R.T. 0.01 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Tgt Ion:142 Resp: 205253
 Ion Ratio Lower Upper
 142 100
 141 91.1 72.6 109.0
 115 39.2 38.2 57.2




Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E
 Ion 115.00 (114.70 to 115.70): E






#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 10.00 min Scan# 692
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:164 Resp: 72575
 Ion Ratio Lower Upper
 164 100
 162 107.2 82.2 123.4
 160 45.5 36.1 54.1



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 43.87 ng
 RT: 9.12 min Scan# 615
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04




Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS


Tgt Ion:	216	Resp:	107654
Ion	Ratio	Lower	Upper
216	100		
214	78.9	62.6	93.8
179	23.5	18.0	27.0
108	23.6	15.9	23.9



#40
 Hexachlorocyclopentadiene
 Concen: 96.35 ng
 RT: 9.10 min Scan# 614
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:	237	Resp:	124524
Ion	Ratio	Lower	Upper
237	100		
235	63.4	47.3	87.3
272	10.2	0.0	31.5









#41
 2,4,6-Tribromophenol
 Concen: 164.08 ng
 RT: 10.78 min Scan# 761
 Delta R.T. 0.01 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Tgt Ion:330 Resp: 94357
 Ion Ratio Lower Upper
 330 100
 332 91.5 78.3 117.5
 141 42.8 37.9 56.9




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#42
 2,4,6-Trichlorophenol
 Concen: 44.63 ng
 RT: 9.22 min Scan# 624
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:196 Resp: 61427
 Ion Ratio Lower Upper
 196 100
 198 93.4 73.9 110.9
 200 28.5 23.4 35.0



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 47.55 ng
 RT: 9.25 min Scan# 627
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Tgt Ion:196 Resp: 66080
 Ion Ratio Lower Upper
 196 100
 198 95.1 76.9 115.3
 97 90.3 63.0 94.4
 132 43.6 30.1 45.1




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 77.51 ng
 RT: 9.31 min Scan# 632
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:172 Resp: 387693
 Ion Ratio Lower Upper
 172 100
 171 36.7 27.7 41.5
 170 25.1 19.1 28.7



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 45.85 ng
 RT: 9.41 min Scan# 641
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Tgt Ion:154 Resp: 270852
 Ion Ratio Lower Upper
 154 100
 153 39.6 21.8 61.8
 76 12.0 0.0 32.1




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0






#46
 2-Chloronaphthalene
 Concen: 44.51 ng
 RT: 9.44 min Scan# 643
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:162 Resp: 212739
 Ion Ratio Lower Upper
 162 100
 127 45.8 37.6 56.4
 164 33.1 27.4 41.0



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E





#47

2-Nitroaniline

Concen: 48.32 ng

RT: 9.53 min Scan# 651

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion: 65 Resp: 70423

Ion Ratio Lower Upper




65 100

92 62.6 48.7 73.1

138 86.4 66.8 100.2



Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): E



#48

Acenaphthylene

Concen: 45.90 ng

RT: 9.86 min Scan# 680

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion: 152 Resp: 324952

Ion Ratio Lower Upper



152 100


151 20.2 16.4 24.6

153 13.2 9.7 14.5



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 52.44 ng
RT: 9.71 min Scan# 667
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:163 Resp: 280022
Ion Ratio Lower Upper
163 100
194 3.3 2.5 3.7
164 10.0 8.1 12.1




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 47.16 ng
RT: 9.77 min Scan# 672
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:165 Resp: 47106
Ion Ratio Lower Upper
165 100
63 87.8 59.0 88.6
89 81.9 56.9 85.3



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 50.07 ng

RT: 10.03 min Scan# 695

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion:154 Resp: 212969

Ion Ratio Lower Upper

154 100

153 111.9 92.2 138.2




152 53.2 43.6 65.4



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 37.44 ng

RT: 9.94 min Scan# 687

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion:138 Resp: 40666

Ion Ratio Lower Upper

138 100

108 12.4 13.7 20.5#



92 108.1 81.0 121.4




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 86.63 ng
RT: 10.04 min Scan# 696
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:184 Resp: 32904
Ion Ratio Lower Upper
184 100
63 61.4 54.9 82.3
154 84.5 94.2 141.4#




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 47.56 ng
RT: 10.20 min Scan# 710
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:168 Resp: 292334
Ion Ratio Lower Upper
168 100
139 38.2 30.7 46.1
169 12.9 9.8 14.8



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 96.81 ng
RT: 10.09 min Scan# 700
Delta R.T. 0.01 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04




Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS


Tgt Ion:139 Resp: 75961
Ion Ratio Lower Upper
139 100
109 80.4 75.9 115.9
65 76.1 109.4 149.4#



#56
2,4-Dinitrotoluene
Concen: 50.28 ng
RT: 10.17 min Scan# 707
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:165 Resp: 62904
Ion Ratio Lower Upper
165 100
63 61.3 48.1 72.1
89 91.3 77.4 116.0
182 2.2 1.3 1.9#





#57

Fluorene

Concen: 46.54 ng

RT: 10.54 min Scan# 740

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion:166 Resp: 229693

Ion Ratio Lower Upper

166 100

165 100.1 77.7 116.5




167 13.2 10.5 15.7



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 51.16 ng

RT: 10.32 min Scan# 720

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Tgt Ion:232 Resp: 55317


Ion Ratio Lower Upper

232 100

131 51.0 41.4 62.0

130 2.7 1.9 2.9

166 31.9 25.7 38.5






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E









#59
Diethylphthalate
Concen: 45.44 ng
RT: 10.41 min Scan# 728
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:149 Resp: 226451
Ion Ratio Lower Upper
149 100
177 17.1 14.9 22.3
150 13.0 10.0 15.0




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 44.43 ng
RT: 10.53 min Scan# 739
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:204 Resp: 98488
Ion Ratio Lower Upper
204 100
206 34.4 27.1 40.7
141 83.7 68.3 102.5



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 45.84 ng

RT: 10.54 min Scan# 740

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion:138 Resp: 50143

Ion Ratio Lower Upper

138 100

92 50.4 22.6 62.6




108 94.0 62.0 102.0



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 48.73 ng

RT: 10.69 min Scan# 753

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Tgt Ion: 77 Resp: 254382


Ion Ratio Lower Upper

77 100

182 19.3 0.0 37.8

105 20.3 0.1 40.1

51 37.3 18.8 58.8






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.48 min Scan# 822
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:188 Resp: 124713
Ion Ratio Lower Upper
188 100
94 5.9 6.1 9.1#
80 6.7 7.0 10.6#




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0






#64
4,6-Dinitro-2-methylphenol
Concen: 52.85 ng
RT: 10.58 min Scan# 743
Delta R.T. 0.01 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:198 Resp: 23153
Ion Ratio Lower Upper
198 100
51 32.4 32.0 72.0
105 32.8 16.3 56.3



Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BF0
Ion 105.00 (104.70 to 105.70): E









#65
 n-Nitrosodiphenylamine
 Concen: 47.10 ng
 RT: 10.65 min Scan# 749
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Tgt Ion:169 Resp: 192753
 Ion Ratio Lower Upper
 169 100
 168 66.8 53.3 79.9
 167 37.8 29.8 44.6




Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E






#66
 4-Bromophenyl-phenylether
 Concen: 47.74 ng
 RT: 11.02 min Scan# 782
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:248 Resp: 60523
 Ion Ratio Lower Upper
 248 100
 250 99.0 75.4 113.2
 141 102.7 93.4 140.0



Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 48.24 ng
RT: 11.09 min Scan# 788
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:284 Resp: 69044
Ion Ratio Lower Upper
284 100
142 23.0 22.4 33.6
249 29.1 25.0 37.4




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 53.42 ng
RT: 11.17 min Scan# 795
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:200 Resp: 60058
Ion Ratio Lower Upper
200 100
173 25.8 4.6 44.6
215 44.2 22.9 62.9



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 101.45 ng
 RT: 11.28 min Scan# 804
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Tgt Ion: 266 Resp: 67760
 Ion Ratio Lower Upper
 266 100
 268 65.6 51.5 77.3
 264 65.3 48.8 73.2




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 48.35 ng
 RT: 11.50 min Scan# 824
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion: 178 Resp: 340372
 Ion Ratio Lower Upper
 178 100
 176 20.2 15.2 22.8
 179 14.9 12.5 18.7



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 48.13 ng
 RT: 11.55 min Scan# 828
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Tgt Ion:178 Resp: 317312
 Ion Ratio Lower Upper
 178 100
 176 18.6 15.1 22.7
 179 15.4 12.0 18.0




Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E






#72
 Carbazole
 Concen: 51.44 ng
 RT: 11.71 min Scan# 842
 Delta R.T. 0.00 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:167 Resp: 293211
 Ion Ratio Lower Upper
 167 100
 166 21.5 18.1 27.1
 139 13.2 10.7 16.1



Abundance Ion 167.00 (166.70 to 167.70): E
 Ion 166.00 (165.70 to 166.70): E
 Ion 139.00 (138.70 to 139.70): E









#73
Di-n-butylphthalate
Concen: 49.22 ng
RT: 12.04 min Scan# 871
Delta R.T. 0.00 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion:149 Resp: 326071
Ion Ratio Lower Upper
149 100
150 8.5 7.2 10.8
104 6.4 3.8 5.8#




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 104.00 (103.70 to 104.70): E






#74
Fluoranthene
Concen: 49.44 ng
RT: 12.70 min Scan# 929
Delta R.T. 0.01 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:202 Resp: 311460
Ion Ratio Lower Upper
202 100
101 5.3 0.0 31.1
203 17.4 0.0 37.5



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.23 min Scan# 1062

Delta R.T. 0.05 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion:240 Resp: 108351

Ion Ratio Lower Upper

240 100

120 10.1 9.9 14.9




236 26.0 19.1 28.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 40.23 ng

RT: 12.82 min Scan# 939

Delta R.T. 0.00 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion:184 Resp: 128093

Ion Ratio Lower Upper

184 100

185 14.7 11.2 16.8



183 12.1 9.6 14.4




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
 Pyrene
 Concen: 45.97 ng
 RT: 12.93 min Scan# 949
 Delta R.T. 0.01 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.8	16.9	25.3
203	17.4	14.5	21.7




Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 200.00 (199.70 to 200.70): E
 Ion 203.00 (202.70 to 203.70): E






#78
 Terphenyl-d14
 Concen: 72.41 ng
 RT: 13.08 min Scan# 962
 Delta R.T. 0.01 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.4	6.0	9.0
122	12.8	8.2	12.2#



Abundance Ion 244.00 (243.70 to 244.70): E
 Ion 212.00 (211.70 to 212.70): E
 Ion 122.00 (121.70 to 122.70): E









#79
Butylbenzylphthalate
Concen: 51.64 ng
RT: 13.60 min Scan# 1007
Delta R.T. 0.02 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04




Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS


Tgt Ion:149 Resp: 136029
Ion Ratio Lower Upper
149 100
91 68.4 53.0 79.6
206 18.1 15.0 22.6



#80
Benzo(a)anthracene
Concen: 45.55 ng
RT: 14.21 min Scan# 1061
Delta R.T. 0.05 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:228 Resp: 278297
Ion Ratio Lower Upper
228 100
226 27.2 22.0 33.0
229 20.2 15.7 23.5









#81
 3,3'-Dichlorobenzidine
 Concen: 43.39 ng
 RT: 14.18 min Scan# 1058
 Delta R.T. 0.05 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04




Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS


Tgt Ion:252 Resp: 75760
 Ion Ratio Lower Upper
 252 100
 254 65.6 49.2 73.8
 126 8.3 8.1 12.1



#82
 Chrysene
 Concen: 44.61 ng
 RT: 14.26 min Scan# 1065
 Delta R.T. 0.05 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:228 Resp: 271127
 Ion Ratio Lower Upper
 228 100
 226 30.4 23.0 34.6
 229 19.8 15.0 22.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 52.19 ng
 RT: 14.23 min Scan# 1062
 Delta R.T. 0.05 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04




Instrument :
 BNA_F
 ClientSampleId :
 SOUTHEAST-FLOOR2-72215MS


Tgt Ion:149 Resp: 185475
 Ion Ratio Lower Upper
 149 100
 167 24.4 20.2 30.2
 279 2.2 2.0 3.0



#84
 Di-n-octyl phthalate
 Concen: 52.05 ng
 RT: 14.95 min Scan# 1125
 Delta R.T. 0.08 min
 Lab File: BF080611.D
 Acq: 24 Jul 2015 1:04

Tgt Ion:149 Resp: 262002
 Ion Ratio Lower Upper
 149 100
 167 1.4 1.2 1.8
 43 12.8 12.5 18.7









#85
Indeno(1,2,3-cd)pyrene
Concen: 61.23 ng
RT: 17.36 min Scan# 1336
Delta R.T. 0.11 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS

Tgt Ion	Ratio	Lower	Upper
276	100		
138	19.9	18.9	28.3
277	24.1	19.9	29.9




Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E






#86
Perylene-d12
Concen: 20.00 ng
RT: 15.85 min Scan# 1204
Delta R.T. 0.10 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.1	17.8	26.8
265	21.5	15.6	23.4



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E









#87
Benzo(b)fluoranthene
Concen: 49.47 ng
RT: 15.41 min Scan# 1166
Delta R.T. 0.10 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04




Instrument :
BNA_F
ClientSampleId :
SOUTHEAST-FLOOR2-72215MS


Tgt Ion:252 Resp: 247828
Ion Ratio Lower Upper
252 100
253 21.6 16.6 25.0
125 5.7 6.1 9.1#



#88
Benzo(k)fluoranthene
Concen: 49.71 ng
RT: 15.41 min Scan# 1166
Delta R.T. 0.07 min
Lab File: BF080611.D
Acq: 24 Jul 2015 1:04

Tgt Ion:252 Resp: 247828
Ion Ratio Lower Upper
252 100
253 21.6 17.0 25.6
125 5.7 8.6 13.0#





#89

Benzo(a)pyrene

Concen: 50.17 ng

RT: 15.79 min Scan# 1199

Delta R.T. 0.10 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS


Tgt Ion:252 Resp: 222963

Ion Ratio Lower Upper

252 100

253 21.7 19.0 28.6




125 7.1 11.2 16.8#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 55.74 ng

RT: 17.38 min Scan# 1338

Delta R.T. 0.11 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04


Tgt Ion:278 Resp: 233962

Ion Ratio Lower Upper

278 100

139 13.8 11.6 17.4



279 22.9 18.9 28.3






Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 59.46 ng

RT: 17.80 min Scan# 1375

Delta R.T. 0.13 min

Lab File: BF080611.D

Acq: 24 Jul 2015 1:04

Instrument :

BNA_F

ClientSampleId :

SOUTHEAST-FLOOR2-72215MS

Tgt Ion: 276 Resp: 254919

Ion Ratio Lower Upper

276 100

277 23.1 18.0 27.0

138 18.0 17.8 26.6

