

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072316\
 Data File : BF089231.D
 Acq On : 23 Jul 2016 11:58
 Operator : UM/SJ
 Sample : H4120-05
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GW-05-5.0-16.0

Quant Time: Jul 25 03:54:44 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 03:29:41 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.58	152	46744	20.00	ng	0.00
21) Naphthalene-d8	7.86	136	196479	20.00	ng	-0.01
38) Acenaphthene-d10	9.61	164	96443	20.00	ng	0.00
63) Phenanthrene-d10	11.08	188	187493	20.00	ng	0.00
75) Chrysene-d12	13.70	240	132857	20.00	ng	-0.01
86) Perylene-d12	15.04	264	96186	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.15	112	198091	64.45	ng	0.00
7) Phenol-d6	6.23	99	177072	44.68	ng	-0.01
23) Nitrobenzene-d5	7.15	82	386426	104.50	ng	0.00
41) 2,4,6-Tribromophenol	10.40	330	107216	123.48	ng	0.00
44) 2-Fluorobiphenyl	8.95	172	644601	92.09	ng	0.00
78) Terphenyl-d14	12.67	244	548835	89.60	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.35	163	31260	4.22	ng	Qvalue # 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072316\
Data File : BF089231.D
Acq On : 23 Jul 2016 11:58
Operator : UM/SJ
Sample : H4120-05
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GW-05-5.0-16.0

Quant Time: Jul 25 03:54:44 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 03:29:41 2016
Response via : Initial Calibration

