

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072316\
 Data File : BF089263.D
 Acq On : 24 Jul 2016 6:35
 Operator : UM/SJ
 Sample : H4151-08
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-06-COMP

Quant Time: Jul 25 05:16:37 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 03:29:41 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.58	152	46767	20.00	ng	0.00
21) Naphthalene-d8	7.86	136	203079	20.00	ng	-0.01
38) Acenaphthene-d10	9.61	164	84251	20.00	ng	0.00
63) Phenanthrene-d10	11.08	188	133753	20.00	ng	0.00
75) Chrysene-d12	13.70	240	91558	20.00	ng	-0.01
86) Perylene-d12	15.05	264	90185	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.16	112	365150	118.75	ng	0.01
7) Phenol-d6	6.24	99	495892	125.05	ng	0.00
23) Nitrobenzene-d5	7.15	82	304550	79.68	ng	0.00
41) 2,4,6-Tribromophenol	10.40	330	75227	99.17	ng	0.00
44) 2-Fluorobiphenyl	8.95	172	503923	82.41	ng	0.00
78) Terphenyl-d14	12.66	244	266238	63.07	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.34	163	76804	11.88	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072316\
Data File : BF089263.D
Acq On : 24 Jul 2016 6:35
Operator : UM/SJ
Sample : H4151-08
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-06-COMP

Quant Time: Jul 25 05:16:37 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 03:29:41 2016
Response via : Initial Calibration

