


Data Path : Z:\HPCHEM1\BNA\_F\Data\BF072316\  
 Data File : BF089219.D  
 Acq On : 23 Jul 2016 4:46  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF072016.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Jul 25 16:56:43 2016



AutoFind: Scans 854, 855, 856; Background Corrected with Scan 851

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 42.0      | 6264    | PASS             |
| 68          | 69           | 0.00         | 2            | 1.9       | 129     | PASS             |
| 69          | 198          | 0.00         | 100          | 44.7      | 6666    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 56.5      | 8435    | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 14929   | PASS             |
| 199         | 198          | 5            | 9            | 6.4       | 957     | PASS             |
| 275         | 198          | 10           | 60           | 24.0      | 3589    | PASS             |
| 365         | 198          | 1            | 100          | 2.6       | 392     | PASS             |
| 441         | 442          | 0.01         | 24           | 16.2      | 1661    | PASS             |
| 442         | 198          | 50           | 100          | 68.7      | 10250   | PASS             |
| 443         | 442          | 15           | 24           | 18.9      | 1936    | PASS             |