

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072321\
 Data File : BF124721.D
 Acq On : 23 Jul 2021 19:39
 Operator : JU/CG
 Sample : M3114-07
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 20210594-AMENDED-COM-8-SPLP

Quant Time: Jul 23 23:29:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 15:28:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	6872	20.00	ng	0.00
21) Naphthalene-d8	8.151	136	28556	20.00	ng	0.00
39) Acenaphthene-d10	9.904	164	15959	20.00	ng	-0.01
64) Phenanthrene-d10	11.392	188	27831	20.00	ng	0.00
76) Chrysene-d12	14.027	240	17289	20.00	ng	-0.01
86) Perylene-d12	15.498	264	13732	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	24892	61.33	ng	0.00
7) Phenol-d6	6.481	99	18408	36.18	ng	-0.02
23) Nitrobenzene-d5	7.434	82	42552	88.69	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	18714	136.58	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	95826	88.19	ng	0.00
79) Terphenyl-d14	12.986	244	100961	100.10	ng	0.00
Target Compounds						
19) n-Nitroso-di-n-propyla...	7.434	70	5519	20.31	ng	Qvalue # 73
25) Isophorone	7.434	82	42552	50.16	ng	# 59
51) 2,6-Dinitrotoluene	9.904	165	1833	7.55	ng	# 26
57) 2,4-Dinitrotoluene	9.904	165	1833	5.56	ng	# 18
67) 4-Bromophenyl-phenylether	10.698	248	1373	4.65	ng	# 1
77) Benzidine	12.986	184	1553	3.19	ng	# 69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072321\
Data File : BF124721.D
Acq On : 23 Jul 2021 19:39
Operator : JU/CG
Sample : M3114-07
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
20210594-AMENDED-COM-8-SPLP

Quant Time: Jul 23 23:29:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 22 15:28:57 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.869 min Scan# 813
Delta R.T. 0.000 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39




Instrument :
BNA_F
ClientSampleId :
20210594-AMENDED-COM-8-SPLP


Tgt Ion:152 Resp: 6872
Ion Ratio Lower Upper
152 100
150 159.9 117.6 176.4
115 61.4 47.1 70.7



#5
2-Fluorophenol
Concen: 61.33 ng
RT: 5.481 min Scan# 577
Delta R.T. 0.000 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39

Tgt Ion:112 Resp: 24892
Ion Ratio Lower Upper
112 100
64 57.4 44.2 66.2
63 30.8 26.2 39.2









#7
Phenol-d6
Concen: 36.18 ng
RT: 6.481 min Scan# 747
Delta R.T. -0.018 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39




Instrument :
BNA_F
ClientSampleId :
20210594-AMENDED-COM-8-SPLP


Tgt Ion:	99	Resp:	18408
Ion	Ratio	Lower	Upper
99	100		
42	25.9	23.5	35.3
71	35.1	25.6	38.4



#19
n-Nitroso-di-n-propylamine
Concen: 20.31 ng
RT: 7.434 min Scan# 909
Delta R.T. 0.141 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39

Tgt Ion:	70	Resp:	5519
Ion	Ratio	Lower	Upper
70	100		
42	72.6	72.2	108.2
101	0.0	7.8	11.8#
130	0.0	21.8	32.6#









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.151 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39




Instrument :
BNA_F
ClientSampleId :
20210594-AMENDED-COM-8-SPLP


Tgt Ion:136	Resp:	28556
Ion Ratio	Lower	Upper
136	100	
137	10.1	9.8 14.6
54	7.9	6.4 9.6
68	3.9	3.3 4.9



#23
Nitrobenzene-d5
Concen: 88.69 ng
RT: 7.434 min Scan# 909
Delta R.T. -0.006 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39

Tgt Ion: 82	Resp:	42552
Ion Ratio	Lower	Upper
82	100	
128	47.9	40.6 60.8
54	60.8	49.6 74.4









#25
Isophorone
Concen: 50.16 ng
RT: 7.434 min Scan# 909
Delta R.T. -0.265 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39




Instrument :
BNA_F
ClientSampleId :
20210594-AMENDED-COM-8-SPLP


Tgt Ion: 82 Resp: 42552
Ion Ratio Lower Upper
82 100
95 0.0 6.5 9.7#
138 0.0 18.4 27.6#



#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.904 min Scan# 1329
Delta R.T. -0.012 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39

Tgt Ion: 164 Resp: 15959
Ion Ratio Lower Upper
164 100
162 100.8 81.3 121.9
160 42.7 38.0 57.0









#42
2,4,6-Tribromophenol
Concen: 136.58 ng
RT: 10.698 min Scan# 1464
Delta R.T. -0.006 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39




Instrument :
BNA_F
ClientSampleId :
20210594-AMENDED-COM-8-SPLP


Tgt Ion: 330 Resp: 18714
Ion Ratio Lower Upper
330 100
332 95.1 76.2 114.4
141 49.3 44.6 66.8



#45
2-Fluorobiphenyl
Concen: 88.19 ng
RT: 9.228 min Scan# 1214
Delta R.T. -0.006 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39

Tgt Ion: 172 Resp: 95826
Ion Ratio Lower Upper
172 100
171 33.4 26.5 39.7
170 23.0 18.8 28.2









#51
2,6-Dinitrotoluene
Concen: 7.55 ng
RT: 9.904 min Scan# 1329
Delta R.T. 0.206 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39




Instrument :
BNA_F
ClientSampleId :
20210594-AMENDED-COM-8-SPLP


Tgt Ion:165 Resp: 1833
Ion Ratio Lower Upper
165 100
63 0.0 42.6 63.8#
89 0.0 40.6 60.8#



#57
2,4-Dinitrotoluene
Concen: 5.56 ng
RT: 9.904 min Scan# 1329
Delta R.T. -0.194 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39

Tgt Ion:165 Resp: 1833
Ion Ratio Lower Upper
165 100
63 0.0 40.9 61.3#
89 0.0 59.7 89.5#
182 0.0 0.0 0.0









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.392 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39




Instrument :
BNA_F
ClientSampleId :
20210594-AMENDED-COM-8-SPLP


Tgt Ion:188 Resp: 27831
Ion Ratio Lower Upper
188 100
94 7.3 6.7 10.1
80 10.1 7.5 11.3



#67
4-Bromophenyl-phenylether
Concen: 4.65 ng
RT: 10.698 min Scan# 1464
Delta R.T. -0.247 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39

Tgt Ion:248 Resp: 1373
Ion Ratio Lower Upper
248 100
250 194.3 74.2 111.4#
141 568.9 60.4 90.6#









#76
Chrysene-d12
Concen: 20.00 ng
RT: 14.027 min Scan# 2030
Delta R.T. -0.012 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39




Instrument :
BNA_F
ClientSampleId :
20210594-AMENDED-COM-8-SPLP


Tgt Ion:240 Resp: 17289
Ion Ratio Lower Upper
240 100
120 12.8 10.8 16.2
236 26.8 21.6 32.4



#77
Benzidine
Concen: 3.19 ng
RT: 12.986 min Scan# 1853
Delta R.T. 0.253 min
Lab File: BF124721.D
Acq: 23 Jul 2021 19:39

Tgt Ion:184 Resp: 1553
Ion Ratio Lower Upper
184 100
185 0.0 10.8 16.2#
183 0.0 8.1 12.1#









#79
 Terphenyl-d14
 Concen: 100.10 ng
 RT: 12.986 min Scan# 1853
 Delta R.T. 0.000 min
 Lab File: BF124721.D
 Acq: 23 Jul 2021 19:39

Instrument :
 BNA_F
 ClientSampleId :
 20210594-AMENDED-COM-8-SPLP

Tgt Ion:	244	Resp:	100961
Ion Ratio	Lower	Upper	
244	100		
212	7.1	5.9	8.9
122	9.9	9.0	13.4



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 15.498 min Scan# 2280
 Delta R.T. -0.012 min
 Lab File: BF124721.D
 Acq: 23 Jul 2021 19:39

Tgt Ion:	264	Resp:	13732
Ion Ratio	Lower	Upper	
264	100		
260	24.0	19.6	29.4
265	23.5	17.2	25.8

