

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072415\
 Data File : BF080665.D
 Acq On : 25 Jul 2015 12:16
 Operator : TP/UM
 Sample : G3054-02RE
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CONC2RE

Manual Integrations
 APPROVED

apatel
 7/27/2015 4:23:58 PM

Quant Time: Jul 27 02:47:53 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 25 04:28:42 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.93	152	47584	20.00	ng	0.00
21) Naphthalene-d8	8.21	136	162097	20.00	ng	0.00
38) Acenaphthene-d10	9.97	164	83628	20.00	ng	0.00
63) Phenanthrene-d10	11.46	188	149966	20.00	ng	0.00
75) Chrysene-d12	14.28	240	213594	20.00	ng	0.13
86) Perylene-d12	15.99	264	216090	20.00	ng	0.25
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	41129	14.55	ng	0.01
7) Phenol-d6	6.54	99	75956	22.50	ng	0.00
23) Nitrobenzene-d5	7.49	82	43061	13.34	ng	0.00
41) 2,4,6-Tribromophenol	10.76	330	2922	3.89	ng	0.00
44) 2-Fluorobiphenyl	9.29	172	80069	14.77	ng	-0.01
78) Terphenyl-d14	13.09	244	97394	9.82	ng	0.03
Target Compounds						
6) Aniline	6.60	93	23021	4.82	ng	Qvalue 98
15) Benzyl Alcohol	7.07	79	66409	23.68	ng	# 82
24) Nitrobenzene	7.50	77	22494	6.99	ng	97
35) Caprolactam	8.60	113	2410	3.75	ng	# 69
49) Dimethylphthalate	9.69	163	20422	3.41	ng	98
62) Azobenzene	10.67	77	18481	3.25	ng	97
70) Phenanthrene	11.48	178	49904	6.03	ng	98
74) Fluoranthene	12.69	202	66170	8.27	ng	95
77) Pyrene	12.93	202	68655	5.13	ng	99
80) Benzo(a)anthracene	14.27	228	46964m	3.75	ng	
82) Chrysene	14.32	228	48470m	4.13	ng	
87) Benzo(b)fluoranthene	15.53	252	67362m	5.14	ng	
89) Benzo(a)pyrene	15.92	252	44854	3.81	ng	# 84
91) Benzo(g,h,i)perylene	17.93	276	28653	2.27	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF072415\
Data File : BF080665.D
Acq On : 25 Jul 2015 12:16
Operator : TP/UM
Sample : G3054-02RE
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CONC2RE

Manual Integrations
APPROVED

apatel
7/27/2015 4:23:58 PM

Quant Time: Jul 27 02:47:53 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 25 04:28:42 2015
Response via : Initial Calibration

