


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072420\
 Data File : BF120996.D
 Acq On : 24 Jul 2020 11:33
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF071620.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 16 14:20:32 2020



AutoFind: Scans 1615, 1616, 1617; Background Corrected with Scan 1610

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.5	39183	PASS
68	69	0.00	2	1.4	504	PASS
69	198	0.00	100	31.1	36348	PASS
70	69	0.00	2	0.3	122	PASS
127	198	10	80	47.1	55072	PASS
197	198	0.00	2	0.5	635	PASS
198	198	100	100	100.0	116834	PASS
199	198	5	9	6.6	7763	PASS
275	198	10	60	28.0	32690	PASS
365	198	1	100	3.6	4177	PASS
441	198	0.01	100	13.5	15715	PASS
442	198	50	100	85.0	99314	PASS
443	442	15	24	19.0	18854	PASS