

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/24/2025	BNA_F	<u>BF143222.D</u>
Compound Name	Response	Retention Time
DDT	368644	13.639
DDD	8567	13.339
DDE	1840	13.004
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10407	379051	2.75