

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072516\
 Data File : BF089277.D
 Acq On : 25 Jul 2016 14:12
 Operator : UM/SJ
 Sample : H4129-11
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 KSB-14-0.2-2.0

Quant Time: Jul 26 12:55:51 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 16:56:43 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.56	152	63044	20.00	ng	-0.01
21) Naphthalene-d8	7.85	136	237074	20.00	ng	0.00
38) Acenaphthene-d10	9.59	164	100586	20.00	ng	-0.01
63) Phenanthrene-d10	11.06	188	169618	20.00	ng	-0.01
75) Chrysene-d12	13.68	240	118301	20.00	ng	-0.01
86) Perylene-d12	15.03	264	95697	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.14	112	357282	86.19	ng	0.00
7) Phenol-d6	6.23	99	511397	95.67	ng	0.00
23) Nitrobenzene-d5	7.13	82	292668	65.59	ng	-0.01
41) 2,4,6-Tribromophenol	10.38	330	70928	78.32	ng	-0.01
44) 2-Fluorobiphenyl	8.92	172	485375	66.49	ng	-0.01
78) Terphenyl-d14	12.65	244	322274	59.09	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	7.87	128	27546	2.15	ng	98
37) 2-Methylnaphthalene	8.56	142	29141	3.57	ng	97
49) Dimethylphthalate	9.32	163	182782	23.68	ng	99
70) Phenanthrene	11.08	178	27887	3.00	ng	99
74) Fluoranthene	12.26	202	29257	2.99	ng	98
77) Pyrene	12.49	202	25980	2.66	ng	97
80) Benzo(a)anthracene	13.67	228	16251	2.16	ng	# 82
87) Benzo(b)fluoranthene	14.67	252	15604m	2.25	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072516\
Data File : BF089277.D
Acq On : 25 Jul 2016 14:12
Operator : UM/SJ
Sample : H4129-11
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
KSB-14-0.2-2.0

Quant Time: Jul 26 12:55:51 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 16:56:43 2016
Response via : Initial Calibration

