

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072516\
 Data File : BF089288.D
 Acq On : 25 Jul 2016 20:43
 Operator : UM/SJ
 Sample : H4207-03
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B2-S

Quant Time: Jul 26 02:03:04 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 16:56:43 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.56	152	48609	20.00	ng	-0.01
21) Naphthalene-d8	7.85	136	191712	20.00	ng	0.00
38) Acenaphthene-d10	9.59	164	85240	20.00	ng	-0.01
63) Phenanthrene-d10	11.06	188	154063	20.00	ng	-0.01
75) Chrysene-d12	13.68	240	96920	20.00	ng	-0.01
86) Perylene-d12	15.02	264	87103	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.14	112	344643	107.83	ng	0.00
7) Phenol-d6	6.22	99	463094	112.36	ng	-0.01
23) Nitrobenzene-d5	7.13	82	283939	78.69	ng	-0.01
41) 2,4,6-Tribromophenol	10.38	330	77565	101.07	ng	-0.01
44) 2-Fluorobiphenyl	8.92	172	450247	72.78	ng	-0.01
78) Terphenyl-d14	12.64	244	300868	67.33	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	9.32	163	156042	23.86	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072516\
Data File : BF089288.D
Acq On : 25 Jul 2016 20:43
Operator : UM/SJ
Sample : H4207-03
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B2-S

Quant Time: Jul 26 02:03:04 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 16:56:43 2016
Response via : Initial Calibration

