



Data Path : Z:\HPCHEM1\BNA F\DATA\BF072517\
Data File : BF097016.D
Acq On : 25 Jul 2017 9:21
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 26 07:14:27 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 14:11:51 2017
Response via : Initial Calibration



TIC: BF097016.D

(76) Benzidine

12.333min (+0.000) 0.00ng

response 443065

Ion	Exp%	Act%
184.00	100	100
185.00	19.40	14.12#
183.00	12.20	11.43
0.00	0.00	0.00