

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072517\
 Data File : BF097062.D
 Acq On : 26 Jul 2017 7:29
 Operator : SJ/JU
 Sample : I4370-05 5X
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-N6-BASE

Manual Integrations
 APPROVED

Sohil
 7/26/2017 5:58:40 PM

Quant Time: Jul 26 07:57:36 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 14:11:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.50	152	119378	20.00	ng	0.00
21) Naphthalene-d8	7.78	136	496286	20.00	ng	0.00
38) Acenaphthene-d10	9.53	164	190886	20.00	ng	0.00
63) Phenanthrene-d10	11.00	188	258812	20.00	ng	0.00
75) Chrysene-d12	13.62	240	223358	20.00	ng	0.00
86) Perylene-d12	14.97	264	161802	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.10	112	157300	19.86	ng	0.00
7) Phenol-d6	6.16	99	192525	21.30	ng	-0.01
23) Nitrobenzene-d5	7.06	82	94725	11.20	ng	-0.01
41) 2,4,6-Tribromophenol	10.31	330	25504	16.91	ng	-0.01
44) 2-Fluorobiphenyl	8.86	172	170567	15.16	ng	-0.01
78) Terphenyl-d14	12.58	244	88321	8.06	ng	0.00
Target Compounds						
11) bis(2-Chloroethyl)ether	6.23	93	112026	13.209	ng	88
13) 1,4-Dichlorobenzene	6.52	146	22008	2.434	ng	92
14) 1,2-Dichlorobenzene	6.67	146	103144	13.063	ng	96
30) 1,2,4-Trichlorobenzene	7.73	180	22585	3.498	ng	96
49) Dimethylphthalate	9.26	163	38801	2.677	ng	99
87) Benzo(b)fluoranthene	14.62	252	37587m	3.864	ng	
91) Benzo(g,h,i)perylene	16.56	276	25282	3.303	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072517\
Data File : BF097062.D
Acq On : 26 Jul 2017 7:29
Operator : SJ/JU
Sample : I4370-05 5X
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-N6-BASE

Manual Integrations
APPROVED

Sohil
7/26/2017 5:58:40 PM

Quant Time: Jul 26 07:57:36 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 14:11:51 2017
Response via : Initial Calibration

