

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072519\
 Data File : BF115754.D
 Acq On : 25 Jul 2019 12:12
 Operator : HP/JU
 Sample : K3887-01DL 10X
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AOC-7(1)DL

Quant Time: Jul 25 13:17:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 12 14:03:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	126008	20.00	ng	-0.02
21) Naphthalene-d8	8.15	136	515593	20.00	ng	-0.02
39) Acenaphthene-d10	9.90	164	297314	20.00	ng	-0.02
64) Phenanthrene-d10	11.39	188	535737	20.00	ng	-0.02
76) Chrysene-d12	14.03	240	316165	20.00	ng	-0.02
87) Perylene-d12	15.50	264	371094	20.00	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	48794	6.33	ng	-0.02
7) Phenol-d6	6.50	99	72199	7.39	ng	-0.02
23) Nitrobenzene-d5	7.42	82	44095	4.97	ng	-0.03
42) 2,4,6-Tribromophenol	10.69	330	19711	5.68	ng	-0.02
45) 2-Fluorobiphenyl	9.22	172	80481	4.74	ng	-0.03
79) Terphenyl-d14	12.97	244	83028	4.85	ng	-0.02


Target Compounds				Qvalue		
25) Isophorone	7.42	82	44094	2.485	ng	# 64
31) Naphthalene	8.17	128	152430	6.104	ng	96
37) 2-Methylnaphthalene	8.86	142	115926	7.004	ng	96
38) 1-Methylnaphthalene	8.96	142	94041	5.982	ng	97
49) Acenaphthylene	9.77	152	294094	10.255	ng	97
51) 2,6-Dinitrotoluene	9.90	165	45813	9.012	ng	# 29
58) Fluorene	10.45	166	122905	6.539	ng	99
71) Phenanthrene	11.42	178	962401	35.046	ng	98
72) Anthracene	11.47	178	281060	9.903	ng	98
75) Fluoranthene	12.61	202	965212	33.261	ng	98
78) Pyrene	12.84	202	1539449	57.471	ng	99
80) Butylbenzylphthalate	13.51	149	258626	22.649	ng	# 25
81) Benzo(a)anthracene	14.02	228	636882	30.577	ng	95
83) Chrysene	14.06	228	586592	28.054	ng	98
86) Indeno(1,2,3-cd)pyrene	16.97	276	232447	13.085	ng	97
88) Benzo(b)fluoranthene	15.07	252	602152	25.199	ng	98
89) Benzo(k)fluoranthene	15.07	252	602152	28.265	ng	98
90) Benzo(a)pyrene	15.44	252	277258	13.500	ng	# 87
91) Dibenzo(a,h)anthracene	17.15	278	80172	4.447	ng	97
92) Benzo(a,h,i)perylene	17.42	276	247849	13.783	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072519\
Data File : BF115754.D
Acq On : 25 Jul 2019 12:12
Operator : HP/JU
Sample : K3887-01DL 10X
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
AOC-7(1)DL

Quant Time: Jul 25 13:17:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 12 14:03:52 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.87 min Scan# 813
 Delta R.T. -0.02 min
 Lab File: BF115754.D
 Acq: 25 Jul 2019 12:12

Instrument :
 BNA_F
 ClientSampleId :
 AOC-7(1)DL

Tot Ion:152 Resp: 126008
 Ion Ratio Lower Upper
 152 100
 150 153.8 129.7 194.5
 115 60.9 48.2 72.4




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
 Concen: 6.334 ng
 RT: 5.49 min Scan# 578
 Delta R.T. -0.02 min
 Lab File: BF115754.D
 Acq: 25 Jul 2019 12:12

Tgt Ion:112 Resp: 48794
 Ion Ratio Lower Upper
 112 100
 64 56.0 46.8 70.2
 63 31.1 24.1 36.1



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BF1
 Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 7.386 ng

RT: 6.50 min Scan# 750

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Instrument :

BNA_F

ClientSampleId :

AOC-7(1)DL





Tot Ion: 99 Resp: 72199

Ion Ratio Lower Upper

99 100

42 16.6 13.6 20.4

71 32.1 28.5 42.7



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.15 min Scan# 1031

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Tgt Ion:136 Resp: 515593




Ion Ratio Lower Upper

136 100

137 11.1 9.0 13.4

54 7.1 5.6 8.4

68 6.3 5.0 7.6









#23
 Nitrobenzene-d5
 Concen: 4.973 ng
 RT: 7.42 min Scan# 907
 Delta R.T. -0.03 min
 Lab File: BF115754.D
 Acq: 25 Jul 2019 12:12




Instrument :
 BNA_F
 ClientSampleID :
 AOC-7(1)DL


Tot Ion	Ratio	Lower	Upper
82	100		
128	41.7	38.6	57.8
54	48.2	35.6	53.4



#25
 Isophorone
 Concen: 2.485 ng
 RT: 7.42 min Scan# 907
 Delta R.T. -0.29 min
 Lab File: BF115754.D
 Acq: 25 Jul 2019 12:12

Tgt Ion	Ratio	Lower	Upper
82	100		
95	0.0	5.5	8.3#
138	0.0	14.9	22.3#





#31

Naphthalene

Concen: 6.104 ng

RT: 8.17 min Scan# 1035

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Instrument :

BNA_F

ClientSampleId :

AOC-7(1)DL


Tot Ion:128 Resp: 152430

Ion Ratio Lower Upper

128 100

129 10.9 9.9 14.9




127 13.2 12.1 18.1



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#37

2-Methylnaphthalene

Concen: 7.004 ng

RT: 8.86 min Scan# 1152

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12


Tgt Ion:142 Resp: 115926

Ion Ratio Lower Upper

142 100

141 87.7 72.0 108.0



115 42.4 29.9 44.9




Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E









#38
1-Methylnaphthalene
Concen: 5.982 ng
RT: 8.96 min Scan# 1169
Delta R.T. -0.02 min
Lab File: BF115754.D
Acq: 25 Jul 2019 12:12

Instrument :
BNA_F
ClientSampleId :
AOC-7(1)DL

Tot Ion:142 Resp: 94041
Ion Ratio Lower Upper
142 100
141 91.2 75.3 112.9
116 4.3 3.3 4.9




Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 116.00 (115.70 to 116.70): E






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.90 min Scan# 1329
Delta R.T. -0.02 min
Lab File: BF115754.D
Acq: 25 Jul 2019 12:12

Tgt Ion:164 Resp: 297314
Ion Ratio Lower Upper
164 100
162 100.5 82.9 124.3
160 45.3 36.8 55.2



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E









#42
 2,4,6-Tribromophenol
 Concen: 5.680 ng
 RT: 10.69 min Scan# 1463
 Delta R.T. -0.02 min
 Lab File: BF115754.D
 Acq: 25 Jul 2019 12:12




Instrument :
 BNA_F
 ClientSampleId :
 AOC-7(1)DL


Tot Ion:330 Resp: 19711
 Ion Ratio Lower Upper
 330 100
 332 94.5 77.4 116.2
 141 45.5 22.4 33.6#



#45
 2-Fluorobiphenyl
 Concen: 4.738 ng
 RT: 9.22 min Scan# 1213
 Delta R.T. -0.03 min
 Lab File: BF115754.D
 Acq: 25 Jul 2019 12:12

Tgt Ion:172 Resp: 80481
 Ion Ratio Lower Upper
 172 100
 171 34.1 30.6 46.0
 170 23.0 20.5 30.7





#49

Acenaphthylene

Concen: 10.255 ng

RT: 9.77 min Scan# 1306

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Instrument :

BNA_F

ClientSampleId :

AOC-7(1)DL


Tot Ion:152 Resp: 294094

Ion Ratio Lower Upper

152 100

151 19.8 17.7 26.5




153 14.5 12.0 18.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#51

2,6-Dinitrotoluene

Concen: 9.012 ng

RT: 9.90 min Scan# 1329

Delta R.T. 0.20 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12


Tgt Ion:165 Resp: 45813

Ion Ratio Lower Upper

165 100

63 2.0 40.6 61.0#



89 2.4 43.9 65.9#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1





#58

Fluorene

Concen: 6.539 ng

RT: 10.45 min Scan# 1422

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Instrument :

BNA_F

ClientSampleId :

AOC-7(1)DL


Tot Ion:166 Resp: 122905

Ion Ratio Lower Upper

166 100

165 99.0 79.8 119.6




167 13.9 11.2 16.8



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.39 min Scan# 1582

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12


Tgt Ion:188 Resp: 535737

Ion Ratio Lower Upper

188 100

94 9.9 7.8 11.6



80 10.5 8.6 13.0




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1





#71

Phenanthrene

Concen: 35.046 ng

RT: 11.42 min Scan# 1586

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Instrument :

BNA_F

ClientSampleId :

AOC-7(1)DL


Tot Ion:178 Resp: 962401

Ion Ratio Lower Upper

178 100

176 20.3 17.2 25.8




179 16.2 13.5 20.3



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Anthracene

Concen: 9.903 ng

RT: 11.47 min Scan# 1595

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12


Tgt Ion:178 Resp: 281060

Ion Ratio Lower Upper

178 100

176 19.0 15.9 23.9



179 15.2 12.7 19.1




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#75

Fluoranthene

Concen: 33.261 ng

RT: 12.61 min Scan# 1789

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Instrument :

BNA_F

ClientSampleId :

AOC-7(1)DL


Tot Ion:202 Resp: 965212

Ion Ratio Lower Upper

202 100

101 12.6 0.0 31.4




203 18.6 0.0 38.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.03 min Scan# 2031

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12


Tgt Ion:240 Resp: 316165

Ion Ratio Lower Upper

240 100

120 11.8 8.6 13.0



236 27.2 21.3 31.9




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#78
Pvrene
Concen: 57.471 ng
RT: 12.84 min Scan# 1828
Delta R.T. -0.02 min
Lab File: BF115754.D
Acq: 25 Jul 2019 12:12

Instrument :
BNA_F
ClientSampleId :
AOC-7(1)DL

Tot Ion:202 Resp: 1539449
Ion Ratio Lower Upper
202 100
200 22.6 18.7 28.1
203 18.5 15.2 22.8




Abundance Ion 202.00 (201.70 to 202.70): E
2000000
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E






#79
Terphenyl-d14
Concen: 4.846 ng
RT: 12.97 min Scan# 1851
Delta R.T. -0.02 min
Lab File: BF115754.D
Acq: 25 Jul 2019 12:12

Tgt Ion:244 Resp: 83028
Ion Ratio Lower Upper
244 100
212 6.5 6.0 9.0
122 11.0 8.6 12.8



Abundance Ion 244.00 (243.70 to 244.70): E
100000
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E









#80
Butylbenzylphthalate
Concen: 22.649 ng
RT: 13.51 min Scan# 1942
Delta R.T. 0.04 min
Lab File: BF115754.D
Acq: 25 Jul 2019 12:12




Instrument :
BNA_F
ClientSampleId :
AOC-7(1)DL


Tot Ion:149 Resp: 258626
Ion Ratio Lower Upper
149 100
91 1.3 55.8 83.8#
206 0.3 18.2 27.2#



#81
Benzo(a)anthracene
Concen: 30.577 ng
RT: 14.02 min Scan# 2029
Delta R.T. -0.02 min
Lab File: BF115754.D
Acq: 25 Jul 2019 12:12

Tgt Ion:228 Resp: 636882
Ion Ratio Lower Upper
228 100
226 33.6 23.4 35.0
229 20.9 17.2 25.8





#83

Chrysene

Concen: 28.054 ng

RT: 14.06 min Scan# 2035

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Instrument :

BNA_F

ClientSampleID :

AOC-7(1)DL





Tot Ion:228 Resp: 586592

Ion Ratio Lower Upper

228 100

226 30.7 25.7 38.5

229 21.1 16.6 25.0



#86

Indeno(1,2,3-cd)pyrene

Concen: 13.085 ng

RT: 16.97 min Scan# 2531

Delta R.T. -0.03 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12




Tgt Ion:276 Resp: 232447


Ion Ratio Lower Upper

276 100

138 23.9 20.2 30.2

277 24.1 20.4 30.6





#87

Pervlene-d12

Concen: 20.000 ng

RT: 15.50 min Scan# 2281

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Instrument :

BNA_F

ClientSampleId :

AOC-7(1)DL


Tot Ion:264 Resp: 371094

Ion Ratio Lower Upper

264 100

260 24.2 19.8 29.8

265 24.1 17.7 26.5






Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88

Benzo(b)fluoranthene

Concen: 25.199 ng

RT: 15.07 min Scan# 2208

Delta R.T. -0.02 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12


Tgt Ion:252 Resp: 602152

Ion Ratio Lower Upper

252 100

253 23.0 18.1 27.1

125 12.9 9.0 13.4






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E






#89
Benzo(k)fluoranthene
Concen: 28.265 ng
RT: 15.07 min Scan# 2208
Delta R.T. -0.05 min
Lab File: BF115754.D
Acq: 25 Jul 2019 12:12

Instrument :
BNA_F
ClientSampleId :
AOC-7(1)DL

Tot Ion:252	Resp:	602152
Ion Ratio	Lower	Upper
252	100	
253	23.0	18.6 28.0
125	12.9	9.0 13.6






Abundance

Ion 252.00 (251.70 to 252.70): E


Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90
Benzo(a)pyrene
Concen: 13.500 ng
RT: 15.44 min Scan# 2270
Delta R.T. -0.02 min
Lab File: BF115754.D
Acq: 25 Jul 2019 12:12

Tgt Ion:252	Resp:	277258
Ion Ratio	Lower	Upper
252	100	
253	24.5	17.8 26.6
125	23.1	9.1 13.7#






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#91

Dibenzo(a,h)anthracene

Concen: 4.447 ng

RT: 17.15 min Scan# 2561

Delta R.T. 0.13 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12

Instrument :

BNA_F

ClientSampleId :

AOC-7(1)DL


Tot Ion:278 Resp: 80172

Ion Ratio Lower Upper

278 100

139 15.2 13.1 19.7




279 26.1 19.6 29.4



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 13.783 ng

RT: 17.42 min Scan# 2607

Delta R.T. -0.03 min

Lab File: BF115754.D

Acq: 25 Jul 2019 12:12


Tgt Ion:276 Resp: 247849

Ion Ratio Lower Upper

276 100

277 23.0 19.0 28.4

138 22.0 17.2 25.8



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

