

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072616\
 Data File : BF089310.D
 Acq On : 26 Jul 2016 13:34
 Operator : UM/SJ
 Sample : H4207-07 5X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NW-2

Manual Integrations
 APPROVED

sohil
 7/27/2016 7:11:58 PM

Quant Time: Jul 26 14:36:54 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 16:56:43 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.56	152	35839	20.00	ng	-0.01
21) Naphthalene-d8	7.84	136	148525	20.00	ng	-0.01
38) Acenaphthene-d10	9.60	164	71154	20.00	ng	0.00
63) Phenanthrene-d10	11.06	188	131152	20.00	ng	-0.01
75) Chrysene-d12	13.68	240	97610	20.00	ng	-0.01
86) Perylene-d12	15.02	264	74631	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.12	112	62668	26.59	ng	-0.02
7) Phenol-d6	6.20	99	84725	27.88	ng	-0.02
23) Nitrobenzene-d5	7.13	82	64509	23.08	ng	-0.01
41) 2,4,6-Tribromophenol	10.38	330	14160	22.10	ng	-0.01
44) 2-Fluorobiphenyl	8.92	172	92306	17.87	ng	-0.01
78) Terphenyl-d14	12.64	244	76352	16.97	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	7.86	128	108174	13.49	ng	# 89
37) 2-Methylnaphthalene	8.56	142	119090m	23.30	ng	
45) 1,1'-Biphenyl	9.03	154	40653	6.34	ng	96
49) Dimethylphthalate	9.32	163	35710	6.54	ng	# 88
51) Acenaphthene	9.62	154	34829	7.69	ng	# 88
57) Fluorene	10.14	166	73262	14.10	ng	# 83
70) Phenanthrene	11.10	178	219297	30.48	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072616\
Data File : BF089310.D
Acq On : 26 Jul 2016 13:34
Operator : UM/SJ
Sample : H4207-07 5X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NW-2

Manual Integrations
APPROVED

sohil
7/27/2016 7:11:58 PM

Quant Time: Jul 26 14:36:54 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 16:56:43 2016
Response via : Initial Calibration

