

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072617\  
 Data File : BF097078.D  
 Acq On : 26 Jul 2017 16:37  
 Operator : SJ/JU  
 Sample : I4443-04  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 E2-M7-NSW

Quant Time: Jul 26 17:57:51 2017  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jul 25 14:11:51 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.49  | 152  | 100791   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 7.77  | 136  | 469212   | 20.00  | ng    | -0.01    |
| 38) Acenaphthene-d10        | 9.52  | 164  | 202771   | 20.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.00 | 188  | 307523   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 13.62 | 240  | 180306   | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 14.97 | 264  | 179949   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.10  | 112  | 693268   | 103.69 | ng    | 0.00     |
| 7) Phenol-d6                | 6.16  | 99   | 859524   | 112.61 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.06  | 82   | 493509   | 61.70  | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol    | 10.31 | 330  | 125528   | 78.35  | ng    | -0.01    |
| 44) 2-Fluorobiphenyl        | 8.86  | 172  | 701410   | 58.70  | ng    | -0.01    |
| 78) Terphenyl-d14           | 12.59 | 244  | 382123   | 43.21  | ng    | 0.00     |
| Target Compounds            |       |      |          |        |       |          |
| 10) Phenol                  | 6.17  | 94   | 47897    | 5.290  | ng    | 90       |
| 11) bis(2-Chloroethyl)ether | 6.24  | 93   | 19499    | 2.723  | ng    | 91       |
| 49) Dimethylphthalate       | 9.26  | 163  | 126195   | 8.195  | ng    | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072617\  
Data File : BF097078.D  
Acq On : 26 Jul 2017 16:37  
Operator : SJ/JU  
Sample : I4443-04  
Misc :  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
E2-M7-NSW

Quant Time: Jul 26 17:57:51 2017  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Jul 25 14:11:51 2017  
Response via : Initial Calibration

