


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072621\
 Data File : BF124757.D
 Acq On : 26 Jul 2021 9:39
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 22 15:28:57 2021






AutoFind: Scans 1625, 1626, 1627; Background Corrected with Scan 1621

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.4	4678	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.2	3274	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.9	4624	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	9864	PASS
199	198	5	9	7.2	708	PASS
275	198	10	60	24.5	2414	PASS
365	198	1	100	2.6	259	PASS
441	198	0.01	100	12.3	1212	PASS
442	442	50	100	100.0	7762	PASS
443	442	15	24	18.4	1428	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072621\
 Data File : BF124757.D
 Acq On : 26 Jul 2021 9:39
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Jul 26 11:08:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 15:28:57 2021
 Response via : Initial Calibration



TIC: BF124757.D\data.ms

(77) Benzidine

12.733min (+ 0.000) 0.00 ng




response 44974

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	13.24
183.00	10.10	11.21
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072621\
Data File : BF124757.D
Acq On : 26 Jul 2021 9:39
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jul 26 11:08:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 22 15:28:57 2021
Response via : Initial Calibration



TIC: BF124757.D\data.ms

(70) Pentachlorophenol (C)

11.198min (+ 0.000) 0.00 ng

response 7840

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	67.65
264.00	69.90	57.38
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/26/2021	BNA_F	BF124757.D
Compound Name	Response	Retention Time
DDT	23811	13.55
DDD	0	13.3
DDE	0	12.71
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
0	23811	0.00