

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072715\
 Data File : BF080687.D
 Acq On : 27 Jul 2015 23:35
 Operator : TP/UM
 Sample : G3088-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DIN46-EFFLUENT-072315

Quant Time: Jul 28 07:51:45 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 28 07:40:31 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.92	152	31997	20.00	ng	0.00
21) Naphthalene-d8	8.21	136	113855	20.00	ng	0.00
38) Acenaphthene-d10	9.96	164	52946	20.00	ng	0.00
63) Phenanthrene-d10	11.45	188	85930	20.00	ng	0.00
75) Chrysene-d12	14.19	240	29641	20.00	ng	-0.06
86) Perylene-d12	15.80	264	14901	20.00	ng	-0.11
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	64851	31.70	ng	0.00
7) Phenol-d6	6.53	99	51435	23.74	ng	0.00
23) Nitrobenzene-d5	7.48	82	199480	89.70	ng	0.00
41) 2,4,6-Tribromophenol	10.75	330	36590	78.15	ng	0.00
44) 2-Fluorobiphenyl	9.29	172	365143	86.31	ng	0.00
78) Terphenyl-d14	13.06	244	201462	115.74	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.68	163	10680	3.05	ng	Qvalue # 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072715\
Data File : BF080687.D
Acq On : 27 Jul 2015 23:35
Operator : TP/UM
Sample : G3088-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DIN46-EFFLUENT-072315

Quant Time: Jul 28 07:51:45 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 28 07:40:31 2015
Response via : Initial Calibration

