


1,4-Dioxane

Response Ratio




Resp Ratio = $5.18 \times 10^{-1} \times \text{Amt} + 9.97 \times 10^{-2}$
Coef of Det (r^2) = 0.995 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
Calibration Table Last Updated: Thu Jul 28 16:02:32 2016

n-Nitrosodimethylamine

Response Ratio




Resp Ratio = 6.01e-001 * Amt - 2.33e-001
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
Calibration Table Last Updated: Thu Jul 28 16:02:32 2016


Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $2.25 \times 10^{-1} \times \text{Amt} - 1.00 \times 10^{-1}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
Calibration Table Last Updated: Thu Jul 28 16:02:32 2016




Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
Calibration Table Last Updated: Thu Jul 28 16:02:32 2016




Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
Calibration Table Last Updated: Thu Jul 28 16:02:32 2016

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.33\text{e-}001 * \text{Amt} - 2.88\text{e-}002$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
Calibration Table Last Updated: Thu Jul 28 16:02:32 2016



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
Calibration Table Last Updated: Thu Jul 28 16:02:32 2016