


Data Path : Z:\HPCHEM1\BNA F\DATA\BF072716\
 Data File : BF089317.D
 Acq On : 27 Jul 2016 8:33
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 28 16:02:32 2016



AutoFind: Scans 851, 852, 853; Background Corrected with Scan 848

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.4	6641	PASS
68	69	0.00	2	1.7	122	PASS
69	198	0.00	100	46.1	7224	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.3	8664	PASS
197	198	0.00	2	0.9	139	PASS
198	198	100	100	100.0	15669	PASS
199	198	5	9	6.2	979	PASS
275	198	10	60	24.1	3778	PASS
365	198	1	100	2.7	419	PASS
441	442	0.01	24	17.1	1877	PASS
442	198	50	100	70.2	10993	PASS
443	442	15	24	19.2	2109	PASS