


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072721\
 Data File : BF124796.D
 Acq On : 27 Jul 2021 16:13
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 22 15:28:57 2021



AutoFind: Scans 1625, 1626, 1627; Background Corrected with Scan 1620

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.9	4684	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.3	3188	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.1	4417	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	9583	PASS
199	198	5	9	6.0	572	PASS
275	198	10	60	26.0	2493	PASS
365	198	1	100	1.8	171	PASS
441	198	0.01	100	13.4	1286	PASS
442	442	50	100	100.0	7716	PASS
443	442	15	24	20.0	1544	PASS