




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072722\
Data File : BF129366.D
Acq On : 27 Jul 2022 09:12
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 28 06:00:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 17:24:17 2022
Response via : Initial Calibration



TIC: BF129366.D\data.ms

(70) Pentachlorophenol (C)




11.216min (-0.006) 563578.59 ng

response 403505

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.70	62.37
264.00	61.80	63.33
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072722\
 Data File : BF129366.D
 Acq On : 27 Jul 2022 09:12
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 28 06:00:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration



TIC: BF129366.D\data.ms

(77) Benzidine**12.757min (+ 0.000) 231293.31 ng****response 2132636**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.83
183.00	11.60	11.91
0.00	0.00	0.00