

Data Path : Z:\HPCHEM1\BNA_F\Data\BF072816\
 Data File : BF089363.D
 Acq On : 28 Jul 2016 13:11
 Operator : UM/SJ
 Sample : H4155-23
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 I-14

Quant Time: Jul 28 13:42:02 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 27 17:12:17 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.55	152	42659	20.00	ng	0.00
21) Naphthalene-d8	7.83	136	197175	20.00	ng	0.00
38) Acenaphthene-d10	9.56	164	90973	20.00	ng	-0.01
63) Phenanthrene-d10	11.04	188	188742	20.00	ng	-0.01
75) Chrysene-d12	13.67	240	133344	20.00	ng	0.00
86) Perylene-d12	15.01	264	96867	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.13	112	350421	131.16	ng	0.02
7) Phenol-d6	6.20	99	472278	133.76	ng	0.00
23) Nitrobenzene-d5	7.12	82	344435	103.11	ng	0.00
41) 2,4,6-Tribromophenol	10.36	330	117278	155.65	ng	0.00
44) 2-Fluorobiphenyl	8.91	172	604982	97.60	ng	0.00
78) Terphenyl-d14	12.63	244	537902	94.41	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.08	88	1175	Below Cal	#	71
15) Benzyl Alcohol	6.70	79	5152	2.35	ng	49
19) n-Nitroso-di-n-propylamine	7.12	70	45172	21.22	ng	77
25) Isophorone	7.12	82	270755	40.24	ng	67
50) 2,6-Dinitrotoluene	9.56	165	11248	7.91	ng	24
55) 4-Nitrophenol	9.79	139	307	6.57	ng	4
56) 2,4-Dinitrotoluene	9.56	165	11248	6.01	ng	7
66) 4-Bromophenyl-phenylether	10.36	248	8356	4.64	ng	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\Data\BF072816\
Data File : BF089363.D
Acq On : 28 Jul 2016 13:11
Operator : UM/SJ
Sample : H4155-23
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
I-14

Quant Time: Jul 28 13:42:02 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF072716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 27 17:12:17 2016
Response via : Initial Calibration

