


Data Path : Z:\HPCHEM1\BNA\_F\Data\BF072816\  
 Data File : BF089340.D  
 Acq On : 28 Jul 2016 1:46  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF072716.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Jul 27 17:12:17 2016



AutoFind: Scans 851, 852, 853; Background Corrected with Scan 848

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 42.0         | 6860       | PASS                |
| 68             | 69              | 0.00            | 2               | 1.6          | 124        | PASS                |
| 69             | 198             | 0.00            | 100             | 46.2         | 7550       | PASS                |
| 70             | 69              | 0.00            | 2               | 0.0          | 0          | PASS                |
| 127            | 198             | 10              | 80              | 57.6         | 9400       | PASS                |
| 197            | 198             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 16330      | PASS                |
| 199            | 198             | 5               | 9               | 6.9          | 1119       | PASS                |
| 275            | 198             | 10              | 60              | 23.3         | 3800       | PASS                |
| 365            | 198             | 1               | 100             | 2.6          | 426        | PASS                |
| 441            | 442             | 0.01            | 24              | 17.1         | 1840       | PASS                |
| 442            | 198             | 50              | 100             | 65.8         | 10740      | PASS                |
| 443            | 442             | 15              | 24              | 19.4         | 2085       | PASS                |