

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072817\
 Data File : BF097164.D
 Acq On : 28 Jul 2017 20:03
 Operator : SJ/JU
 Sample : I4465-04 5X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OK-03-072617-B

Quant Time: Jul 28 23:13:45 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 14:11:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.49	152	118571	20.00	ng	-0.01
21) Naphthalene-d8	7.77	136	512354	20.00	ng	-0.02
38) Acenaphthene-d10	9.52	164	212921	20.00	ng	-0.02
63) Phenanthrene-d10	10.99	188	302038	20.00	ng	-0.02
75) Chrysene-d12	13.61	240	221358	20.00	ng	-0.02
86) Perylene-d12	14.96	264	181049	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.08	112	74626	9.49	ng	-0.02
7) Phenol-d6	6.16	99	101626	11.32	ng	-0.01
23) Nitrobenzene-d5	7.05	82	60147	6.89	ng	-0.02
41) 2,4,6-Tribromophenol	10.30	330	13819	8.21	ng	-0.02
44) 2-Fluorobiphenyl	8.85	172	103639	8.26	ng	-0.02
78) Terphenyl-d14	12.57	244	60355	5.56	ng	-0.02
Target Compounds						Qvalue
59) Diethylphthalate	9.95	149	38568	2.600	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF072817\
Data File : BF097164.D
Acq On : 28 Jul 2017 20:03
Operator : SJ/JU
Sample : I4465-04 5X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OK-03-072617-B

Quant Time: Jul 28 23:13:45 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 14:11:51 2017
Response via : Initial Calibration

