

Data Path : Z:\HPCHEM1\BNA F\Data\BF072817\
 Data File : BF097142.D
 Acq On : 28 Jul 2017 9:35
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 28 13:23:04 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 14:11:51 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.554	0.530	4.3	82	0.00
3	Pyridine	1.695	1.649	2.7	69	0.02
4	n-Nitrosodimethylamine	0.939	0.925	1.5	76	0.00
5 S	2-Fluorophenol	1.327	1.302	1.9	80	0.00
6	Aniline	1.906	2.013	-5.6	82	0.00
7 S	Phenol-d6	1.515	1.620	-6.9	84	0.01
8	2-Chlorophenol	1.419	1.469	-3.5	82	0.01
9	Benzaldehyde	0.897	1.044	-16.4	94	0.01
10 C	Phenol	1.797	1.952	-8.6	83	0.01
11	bis(2-Chloroethyl)ether	1.421	1.430	-0.6	78	0.00
12	1,3-Dichlorobenzene	1.529	1.507	1.4	78	0.01
13 C	1,4-Dichlorobenzene	1.515	1.532	-1.1	85	0.00
14	1,2-Dichlorobenzene	1.323	1.321	0.2	82	0.01
15	Benzyl Alcohol	0.959	0.954	0.5	85	0.01
16	2,2'-oxybis(1-Chloropropane	2.850	2.700	5.3	80	0.00
17	2-Methylphenol	1.095	1.065	2.7	81	0.01
18	Hexachloroethane	0.554	0.569	-2.7	84	0.01
19 P	n-Nitroso-di-n-propylamine	0.997	0.965	3.2	82	0.00
20	3+4-Methylphenols	1.430	1.459	-2.0	86	0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	82	0.01
22	Acetophenone	0.480	0.452	5.8	80	0.00
23 S	Nitrobenzene-d5	0.341	0.330	3.2	83	0.01
24	Nitrobenzene	0.355	0.337	5.1	79	0.01
25	Isophorone	0.668	0.590	11.7	76	0.01
26 C	2-Nitrophenol	0.192	0.196	-2.1	84	0.01
27	2,4-Dimethylphenol	0.317	0.305	3.8	76	0.00
28	bis(2-Chloroethoxy)methane	0.436	0.396	9.2	75	0.00
29 C	2,4-Dichlorophenol	0.273	0.273	0.0	80	0.00
30	1,2,4-Trichlorobenzene	0.260	0.262	-0.8	85	0.01
31	Naphthalene	0.943	0.934	1.0	86	0.01
32	Benzoic acid	0.237	0.226	4.6	74	0.00
33	4-Chloroaniline	0.384	0.392	-2.1	82	0.01
34 C	Hexachlorobutadiene	0.138	0.132	4.3	80	0.01
35	Caprolactam	0.088	0.087	1.1	79	0.01
36 C	4-Chloro-3-methylphenol	0.298	0.289	3.0	78	0.00
37	2-Methylnaphthalene	0.597	0.643	-7.7	89	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.01
39	1,2,4,5-Tetrachlorobenzene	0.534	0.546	-2.2	85	0.01
40 P	Hexachlorocyclopentadiene	0.156	0.168	-7.7	82	0.00
41 S	2,4,6-Tribromophenol	0.158	0.161	-1.9	90	0.00
42 C	2,4,6-Trichlorophenol	0.387	0.392	-1.3	84	0.00
43	2,4,5-Trichlorophenol	0.387	0.395	-2.1	84	0.01
44 S	2-Fluorobiphenyl	1.178	1.185	-0.6	85	0.01

Data Path : Z:\HPCHEM1\BNA F\Data\BF072817\
 Data File : BF097142.D
 Acq On : 28 Jul 2017 9:35
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 28 13:23:04 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 14:11:51 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.708	1.676	1.9	82	0.00
46	2-Chloronaphthalene	1.257	1.227	2.4	82	0.01
47	2-Nitroaniline	0.456	0.442	3.1	76	0.00
48	Acenaphthylene	1.930	1.783	7.6	84	0.00
49	Dimethylphthalate	1.519	1.453	4.3	86	0.00
50	2,6-Dinitrotoluene	0.322	0.316	1.9	86	0.00
51 C	Acenaphthene	1.137	1.070	5.9	84	0.01
52	3-Nitroaniline	0.400	0.383	4.3	86	0.00
53 P	2,4-Dinitrophenol	0.146	0.146	0.0	81	0.00
54	Dibenzofuran	1.514	1.424	5.9	84	0.01
55 P	4-Nitrophenol	0.238	0.209	12.2	72	0.00
56	2,4-Dinitrotoluene	0.370	0.361	2.4	87	0.00
57	Fluorene	1.138	1.183	-4.0	91	0.00
58	2,3,4,6-Tetrachlorophenol	0.267	0.252	5.6	82	0.01
59	Diethylphthalate	1.393	1.411	-1.3	91	0.00
60	4-Chlorophenyl-phenylether	0.470	0.489	-4.0	91	0.00
61	4-Nitroaniline	0.380	0.404	-6.3	91	0.00
62	Azobenzene	1.293	1.257	2.8	79	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
64	4,6-Dinitro-2-methylphenol	0.124	0.131	-5.6	90	0.00
65 c	n-Nitrosodiphenylamine	0.696	0.633	9.1	88	0.00
66	4-Bromophenyl-phenylether	0.200	0.185	7.5	88	0.00
67	Hexachlorobenzene	0.224	0.204	8.9	87	0.00
68	Atrazine	0.200	0.187	6.5	91	0.00
69 C	Pentachlorophenol	0.093	0.083	10.8	77	0.00
70	Phenanthrene	1.072	1.007	6.1	88	0.00
71	Anthracene	1.098	1.016	7.5	87	0.00
72	Carbazole	1.079	1.026	4.9	92	0.00
73	Di-n-butylphthalate	1.334	1.232	7.6	87	0.00
74 C	Fluoranthene	1.050	0.981	6.6	92	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
76	Benzidine	0.742	0.621	16.3	69	0.00
77	Pyrene	1.637	1.604	2.0	93	0.00
78 S	Terphenyl-d14	0.981	0.947	3.5	92	0.00
79	Butylbenzylphthalate	0.833	0.826	0.8	90	0.00
80	Benzo(a)anthracene	1.294	1.260	2.6	91	0.00
81	3,3'-Dichlorobenzidine	0.488	0.471	3.5	90	0.00
82	Chrysene	1.264	1.240	1.9	92	0.00
83	Bis(2-ethylhexyl)phthalate	1.087	1.006	7.5	86	0.00
84 c	Di-n-octyl phthalate	1.996	1.804	9.6	82	0.00
85	Indeno(1,2,3-cd)pyrene	1.084	1.035	4.5	95	0.00
86 I	Perylene-d12	1.000	1.000	0.0	87	0.00
87	Benzo(b)fluoranthene	1.202	1.205	-0.2	91	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF072817\
Data File : BF097142.D
Acq On : 28 Jul 2017 9:35
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 28 13:23:04 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 14:11:51 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.146	1.076	6.1	86	0.00
89 C	Benzo(a)pyrene	1.100	1.099	0.1	93	0.00
90	Dibenzo(a,h)anthracene	0.902	0.900	0.2	95	0.00
91	Benzo(a,h,i)perylene	0.946	0.955	-1.0	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0