

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072820\
 Data File : BF121088.D
 Acq On : 29 Jul 2020 00:59
 Operator : JU/CG
 Sample : L3430-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TS1

Manual Integrations
 APPROVED

mohammad
 7/29/2020 11:47:53 AM

Quant Time: Jul 29 04:23:11 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 16 14:20:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	177213	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	669067	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	369856	20.00	ng	0.00
64) Phenanthrene-d10	11.33	188	718287	20.00	ng	0.00
76) Chrysene-d12	13.97	240	591716	20.00	ng	0.00
86) Perylene-d12	15.43	264	501159	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	966300	89.39	ng	0.02
7) Phenol-d6	6.45	99	1166430	83.53	ng	0.00
23) Nitrobenzene-d5	7.37	82	733295	63.42	ng	0.00
42) 2,4,6-Tribromophenol	10.64	330	394568	97.46	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	1432056	57.81	ng	0.00
79) Terphenyl-d14	12.92	244	1526394	56.08	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.56	163	211875	7.941	ng	99
71) Phenanthrene	11.36	178	92231	2.497	ng	99
75) Fluoranthene	12.54	202	210866	5.150	ng	98
78) Pyrene	12.77	202	213927	5.114	ng	98
81) Benzo(a)anthracene	13.96	228	96751	2.700	ng	97
83) Chrysene	14.00	228	106786	2.836	ng	98
88) Benzo(b)fluoranthene	15.00	252	122942m	4.058	ng	
90) Benzo(a)pyrene	15.36	252	81831	2.961	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072820\
Data File : BF121088.D
Acq On : 29 Jul 2020 00:59
Operator : JU/CG
Sample : L3430-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TS1

Manual Integrations
APPROVED

mohammad
7/29/2020 11:47:53 AM

Quant Time: Jul 29 04:23:11 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 16 14:20:32 2020
Response via : Initial Calibration

