

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
 Data File : BF129408.D
 Acq On : 28 Jul 2022 14:41
 Operator : CG\JU
 Sample : N3894-07
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RVE-26

Quant Time: Jul 29 02:07:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	180193	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	698725	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	344640	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	472703	20.000	ng	0.00
76) Chrysene-d12	14.057	240	314112	20.000	ng	-0.01
86) Perylene-d12	15.551	264	395421	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	1140824	103.134	ng	0.01
7) Phenol-d6	6.522	99	1423947	102.415	ng	0.00
23) Nitrobenzene-d5	7.445	82	904605	70.673	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	295237	89.102	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1568488	70.403	ng	0.00
79) Terphenyl-d14	12.998	244	1061022	63.726	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
Data File : BF129408.D
Acq On : 28 Jul 2022 14:41
Operator : CG\JU
Sample : N3894-07
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RVE-26

Quant Time: Jul 29 02:07:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 17:24:17 2022
Response via : Initial Calibration

