

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
 Data File : BF129411.D
 Acq On : 28 Jul 2022 16:22
 Operator : CG\JU
 Sample : PB146473BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB146473BL

Quant Time: Jul 29 02:08:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	233102	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	950830	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	512508	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	872297	20.000	ng	0.00
76) Chrysene-d12	14.063	240	673845	20.000	ng	0.00
86) Perylene-d12	15.551	264	581386	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1810361	126.514	ng	0.02
7) Phenol-d6	6.522	99	2332627	129.690	ng	0.00
23) Nitrobenzene-d5	7.451	82	1504150	86.356	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	626721	127.192	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	2594082	78.300	ng	0.00
79) Terphenyl-d14	13.010	244	2844446	79.637	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
Data File : BF129411.D
Acq On : 28 Jul 2022 16:22
Operator : CG\JU
Sample : PB146473BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB146473BL

Quant Time: Jul 29 02:08:14 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 17:24:17 2022
Response via : Initial Calibration

