

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
 Data File : BF129399.D
 Acq On : 28 Jul 2022 09:48
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 29 02:06:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	85	0.00
2	1,4-Dioxane	40.000	38.894	2.8	83	0.02
3	Pyridine	40.000	37.403	6.5	78	0.02
4	n-Nitrosodimethylamine	40.000	40.091	-0.2	85	0.02
5 S	2-Fluorophenol	80.000	77.826	2.7	85	0.01
6	Aniline	40.000	37.553	6.1	81	0.00
7 S	Phenol-d6	80.000	77.877	2.7	85	0.01
8	2-Chlorophenol	40.000	39.063	2.3	84	0.00
9	Benzaldehyde	40.000	35.442	11.4	81	0.00
10 C	Phenol	40.000	38.981	2.5	85	0.01
11	bis(2-Chloroethyl)ether	40.000	38.746	3.1	84	0.00
12	1,3-Dichlorobenzene	40.000	39.353	1.6	86	0.00
13 C	1,4-Dichlorobenzene	40.000	39.067	2.3	85	0.00
14	1,2-Dichlorobenzene	40.000	38.581	3.5	83	0.00
15	Benzyl Alcohol	40.000	39.475	1.3	84	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.217	9.5	78	0.00
17	2-Methylphenol	40.000	38.449	3.9	84	0.01
18	Hexachloroethane	40.000	40.051	-0.1	87	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.288	6.8	82	0.00
20	3+4-Methylphenols	40.000	36.606	8.5	83	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	82	0.00
22	Acetophenone	40.000	38.788	3.0	84	0.00
23 S	Nitrobenzene-d5	80.000	80.807	-1.0	84	0.00
24	Nitrobenzene	40.000	39.838	0.4	82	0.00
25	Isophorone	40.000	38.586	3.5	82	0.00
26 C	2-Nitrophenol	40.000	40.352	-0.9	83	0.00
27	2,4-Dimethylphenol	40.000	39.672	0.8	83	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.951	2.6	82	0.00
29 C	2,4-Dichlorophenol	40.000	40.343	-0.9	83	0.00
30	1,2,4-Trichlorobenzene	40.000	40.184	-0.5	84	0.00
31	Naphthalene	40.000	39.697	0.8	84	0.00
32	Benzoic acid	40.000	31.031	22.4	62	0.00
33	4-Chloroaniline	40.000	39.335	1.7	83	0.00
34 C	Hexachlorobutadiene	40.000	42.053	-5.1	89	0.00
35	Caprolactam	40.000	39.268	1.8	82	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.441	-1.1	85	0.00
37	2-Methylnaphthalene	40.000	39.528	1.2	84	0.00
38	1-Methylnaphthalene	40.000	39.704	0.7	85	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	83	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.366	-0.9	86	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.050	-0.1	77	0.00
42 S	2,4,6-Tribromophenol	80.000	80.727	-0.9	86	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.601	1.0	82	0.00
44	2,4,5-Trichlorophenol	40.000	39.677	0.8	84	0.00
45 S	2-Fluorobiphenyl	80.000	74.044	7.4	87	0.00
46	1,1'-Biphenyl	40.000	39.977	0.1	85	0.00
47	2-Chloronaphthalene	40.000	39.482	1.3	84	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
 Data File : BF129399.D
 Acq On : 28 Jul 2022 09:48
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 29 02:06:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.011	-0.0	83	0.00
49	Acenaphthylene	40.000	39.609	1.0	85	0.00
50	Dimethylphthalate	40.000	39.215	2.0	84	0.00
51	2,6-Dinitrotoluene	40.000	39.585	1.0	83	0.00
52 C	Acenaphthene	40.000	39.246	1.9	82	0.00
53	3-Nitroaniline	40.000	38.466	3.8	79	0.00
54 P	2,4-Dinitrophenol	40.000	37.424	6.4	74	0.00
55	Dibenzofuran	40.000	39.705	0.7	84	0.00
56 P	4-Nitrophenol	40.000	36.082	9.8	74	0.01
57	2,4-Dinitrotoluene	40.000	40.657	-1.6	83	0.00
58	Fluorene	40.000	39.257	1.9	85	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.003	2.5	80	0.00
60	Diethylphthalate	40.000	39.477	1.3	84	0.00
61	4-Chlorophenyl-phenylether	40.000	39.303	1.7	85	0.00
62	4-Nitroaniline	40.000	39.685	0.8	83	0.00
63	Azobenzene	40.000	38.872	2.8	82	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	83	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.846	0.4	78	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.017	2.5	83	0.00
67	4-Bromophenyl-phenylether	40.000	40.164	-0.4	85	0.00
68	Hexachlorobenzene	40.000	41.077	-2.7	86	0.00
69	Atrazine	40.000	39.507	1.2	83	0.00
70 C	Pentachlorophenol	40.000	34.422	13.9	68	0.00
71	Phenanthrene	40.000	39.082	2.3	83	0.00
72	Anthracene	40.000	39.287	1.8	85	0.00
73	Carbazole	40.000	38.961	2.6	83	0.00
74	Di-n-butylphthalate	40.000	39.296	1.8	84	0.00
75 C	Fluoranthene	40.000	40.252	-0.6	86	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	88	0.00
77	Benzidine	40.000	37.203	7.0	78	0.00
78	Pyrene	40.000	38.568	3.6	85	0.00
79 S	Terphenyl-d14	80.000	77.907	2.6	88	0.00
80	Butylbenzylphthalate	40.000	39.780	0.5	86	0.00
81	Benzo(a)anthracene	40.000	39.610	1.0	89	0.00
82	3,3'-Dichlorobenzidine	40.000	38.658	3.4	85	0.00
83	Chrysene	40.000	39.500	1.3	90	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.719	3.2	86	0.00
85 c	Di-n-octyl phthalate	40.000	39.212	2.0	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	83	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.505	1.2	76	0.00
88	Benzo(b)fluoranthene	40.000	42.024	-5.1	91	0.00
89	Benzo(k)fluoranthene	40.000	38.935	2.7	80	0.00
90 C	Benzo(a)pyrene	40.000	39.187	2.0	82	0.00
91	Dibenzo(a,h)anthracene	40.000	38.956	2.6	75	0.00
92	Benzo(g,h,i)perylene	40.000	40.004	-0.0	75	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
Data File : BF129399.D
Acq On : 28 Jul 2022 09:48
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 29 02:06:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 17:24:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0