

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
 Data File : BF129410.D
 Acq On : 28 Jul 2022 15:50
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 29 02:08:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	87	0.00
2	1,4-Dioxane	40.000	39.048	2.4	85	0.01
3	Pyridine	40.000	38.138	4.7	81	0.02
4	n-Nitrosodimethylamine	40.000	39.738	0.7	86	0.02
5 S	2-Fluorophenol	80.000	78.062	2.4	87	0.01
6	Aniline	40.000	37.424	6.4	82	0.00
7 S	Phenol-d6	80.000	78.676	1.7	88	0.01
8	2-Chlorophenol	40.000	39.031	2.4	85	0.00
9	Benzaldehyde	40.000	36.041	9.9	84	0.00
10 C	Phenol	40.000	39.658	0.9	88	0.01
11	bis(2-Chloroethyl)ether	40.000	39.008	2.5	86	0.00
12	1,3-Dichlorobenzene	40.000	39.424	1.4	87	0.00
13 C	1,4-Dichlorobenzene	40.000	39.310	1.7	87	0.00
14	1,2-Dichlorobenzene	40.000	38.863	2.8	85	0.00
15	Benzyl Alcohol	40.000	40.223	-0.6	87	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.702	8.2	81	0.00
17	2-Methylphenol	40.000	38.826	2.9	86	0.01
18	Hexachloroethane	40.000	39.234	1.9	86	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.132	4.7	86	0.00
20	3+4-Methylphenols	40.000	36.830	7.9	85	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	85	0.00
22	Acetophenone	40.000	38.945	2.6	87	0.00
23 S	Nitrobenzene-d5	80.000	80.389	-0.5	87	0.00
24	Nitrobenzene	40.000	39.623	0.9	85	0.00
25	Isophorone	40.000	38.973	2.6	85	0.00
26 C	2-Nitrophenol	40.000	40.807	-2.0	86	0.00
27	2,4-Dimethylphenol	40.000	40.011	-0.0	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.523	3.7	84	0.00
29 C	2,4-Dichlorophenol	40.000	39.925	0.2	85	0.00
30	1,2,4-Trichlorobenzene	40.000	39.625	0.9	86	0.00
31	Naphthalene	40.000	39.240	1.9	86	0.00
32	Benzoic acid	40.000	29.881	25.3#	62	0.00
33	4-Chloroaniline	40.000	38.931	2.7	85	0.00
34 C	Hexachlorobutadiene	40.000	40.752	-1.9	89	0.00
35	Caprolactam	40.000	39.814	0.5	86	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.128	-0.3	87	0.00
37	2-Methylnaphthalene	40.000	39.315	1.7	86	0.00
38	1-Methylnaphthalene	40.000	39.667	0.8	88	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.297	-0.7	88	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.144	-0.4	80	0.00
42 S	2,4,6-Tribromophenol	80.000	80.736	-0.9	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.975	0.1	85	0.00
44	2,4,5-Trichlorophenol	40.000	40.093	-0.2	87	0.00
45 S	2-Fluorobiphenyl	80.000	73.854	7.7	89	0.00
46	1,1'-Biphenyl	40.000	39.689	0.8	86	0.00
47	2-Chloronaphthalene	40.000	39.529	1.2	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
 Data File : BF129410.D
 Acq On : 28 Jul 2022 15:50
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 29 02:08:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.602	-1.5	86	0.00
49	Acenaphthylene	40.000	39.543	1.1	87	0.00
50	Dimethylphthalate	40.000	39.210	2.0	86	0.00
51	2,6-Dinitrotoluene	40.000	40.032	-0.1	86	0.00
52 C	Acenaphthene	40.000	39.399	1.5	85	0.00
53	3-Nitroaniline	40.000	38.276	4.3	81	0.00
54 P	2,4-Dinitrophenol	40.000	38.332	4.2	78	0.00
55	Dibenzofuran	40.000	39.506	1.2	86	0.00
56 P	4-Nitrophenol	40.000	36.765	8.1	77	0.01
57	2,4-Dinitrotoluene	40.000	40.444	-1.1	84	0.00
58	Fluorene	40.000	39.339	1.7	88	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.538	1.2	84	0.00
60	Diethylphthalate	40.000	39.106	2.2	85	0.00
61	4-Chlorophenyl-phenylether	40.000	39.085	2.3	87	0.00
62	4-Nitroaniline	40.000	39.323	1.7	84	0.00
63	Azobenzene	40.000	38.706	3.2	84	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	85	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.838	-2.1	82	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.633	3.4	84	0.00
67	4-Bromophenyl-phenylether	40.000	39.872	0.3	86	0.00
68	Hexachlorobenzene	40.000	40.124	-0.3	87	0.00
69	Atrazine	40.000	38.844	2.9	84	0.00
70 C	Pentachlorophenol	40.000	34.795	13.0	71	0.00
71	Phenanthrene	40.000	38.861	2.8	85	0.00
72	Anthracene	40.000	39.201	2.0	87	0.00
73	Carbazole	40.000	38.780	3.0	85	0.00
74	Di-n-butylphthalate	40.000	38.398	4.0	84	0.00
75 C	Fluoranthene	40.000	39.470	1.3	87	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	88	0.00
77	Benzidine	40.000	35.338	11.7	75	0.00
78	Pyrene	40.000	39.756	0.6	88	0.00
79 S	Terphenyl-d14	80.000	79.525	0.6	90	0.00
80	Butylbenzylphthalate	40.000	40.182	-0.5	87	0.00
81	Benzo(a)anthracene	40.000	39.779	0.6	89	0.00
82	3,3'-Dichlorobenzidine	40.000	38.191	4.5	84	0.00
83	Chrysene	40.000	39.661	0.8	90	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.126	2.2	87	0.00
85 c	Di-n-octyl phthalate	40.000	40.057	-0.1	90	0.00
86 I	Perylene-d12	20.000	20.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.291	1.8	80	0.00
88	Benzo(b)fluoranthene	40.000	41.736	-4.3	96	0.00
89	Benzo(k)fluoranthene	40.000	38.553	3.6	84	0.00
90 C	Benzo(a)pyrene	40.000	39.750	0.6	88	0.00
91	Dibenzo(a,h)anthracene	40.000	39.199	2.0	79	0.00
92	Benzo(g,h,i)perylene	40.000	39.502	1.2	78	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072822\
Data File : BF129410.D
Acq On : 28 Jul 2022 15:50
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 29 02:08:02 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 17:24:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0