

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072922\
 Data File : BF129458.D
 Acq On : 30 Jul 2022 00:02
 Operator : CG\JU
 Sample : N3948-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MHA

Quant Time: Jul 30 01:12:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	158942	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	588197	20.000	ng	# 0.00
39) Acenaphthene-d10	9.934	164	290866	20.000	ng	0.00
64) Phenanthrene-d10	11.422	188	426122	20.000	ng	# 0.00
76) Chrysene-d12	14.063	240	312400	20.000	ng	0.00
86) Perylene-d12	15.557	264	337019	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1254221	128.545	ng	0.02
7) Phenol-d6	6.528	99	1575172	128.439	ng	0.01
23) Nitrobenzene-d5	7.451	82	959140	89.015	ng	0.00
42) 2,4,6-Tribromophenol	10.728	330	312707	111.822	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	1513892	80.515	ng	0.00
79) Terphenyl-d14	13.004	244	1185649	71.601	ng	0.00
Target Compounds						Qvalue
18) Hexachloroethane	7.416	117	47155	10.771	ng	# 1
19) n-Nitroso-di-n-propyla...	7.304	70	24024	3.236	ng	# 62
22) Acetophenone	7.275	105	35148	2.567	ng	# 56
25) Isophorone	7.687	82	73650	3.813	ng	# 1
31) Naphthalene	8.192	128	288141	9.642	ng	# 92
35) Caprolactam	8.616	113	14743	5.351	ng	# 1
37) 2-Methylnaphthalene	8.887	142	1094813	56.375	ng	100
38) 1-Methylnaphthalene	8.987	142	719374	38.372	ng	100
46) 1,1'-Biphenyl	9.345	154	58542	2.758	ng	96
52) Acenaphthene	9.963	154	65110	3.823	ng	# 77
53) 3-Nitroaniline	9.881	138	14213	2.678	ng	# 24
55) Dibenzofuran	10.139	168	71303	3.037	ng	# 1
56) 4-Nitrophenol	10.034	139	57769	14.755	ng	# 51
57) 2,4-Dinitrotoluene	10.134	165	32100	5.468	ng	# 56
58) Fluorene	10.481	166	138429	7.369	ng	# 87
66) n-Nitrosodiphenylamine	10.592	169	193128	13.609	ng	# 53
71) Phenanthrene	11.445	178	448405	19.277	ng	97
72) Anthracene	11.445	178	448405	19.616	ng	97
75) Fluoranthene	12.633	202	243194	10.319	ng	98
78) Pyrene	12.863	202	226242	8.660	ng	98
81) Benzo(a)anthracene	14.051	228	120384	5.641	ng	98
83) Chrysene	14.086	228	95664	4.757	ng	95
87) Indeno(1,2,3-cd)pyrene	17.062	276	48645	2.143	ng	95
88) Benzo(b)fluoranthene	15.110	252	185981	8.319	ng	# 97
89) Benzo(k)fluoranthene	15.110	252	185981	8.489	ng	# 97
90) Benzo(a)pyrene	15.486	252	98566	5.431	ng	# 96
92) Benzo(g,h,i)perylene	17.521	276	42830	2.269	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072922\
Data File : BF129458.D
Acq On : 30 Jul 2022 00:02
Operator : CG\JU
Sample : N3948-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MHA

Quant Time: Jul 30 01:12:04 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 17:24:17 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.887 min Scan# 71
 Delta R.T. 0.001 min
 Lab File: BF129458.D
 Acq: 30 Jul 2022 00:02




Instrument :
 BNA_F
 ClientSampleId :
 MHA


Tgt Ion:152 Resp: 158942
 Ion Ratio Lower Upper
 152 100
 150 153.5 123.7 185.5
 115 59.9 49.9 74.9



#5
 2-Fluorophenol
 Concen: 128.545 ng
 RT: 5.522 min Scan# 550
 Delta R.T. 0.024 min
 Lab File: BF129458.D
 Acq: 30 Jul 2022 00:02

Tgt Ion:112 Resp: 1254221
 Ion Ratio Lower Upper
 112 100
 64 61.7 53.3 79.9
 63 31.9 28.0 42.0









#7
Phenol-d6
Concen: 128.439 ng
RT: 6.528 min Scan# 719
Delta R.T. 0.012 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Instrument :
BNA_F
ClientSampleId :
MHA

Tgt Ion: 99 Resp: 1575172




Ion	Ratio	Lower	Upper
99	100		
42	19.4	17.3	25.9
71	35.4	28.9	43.3




#18
Hexachloroethane
Concen: 10.771 ng
RT: 7.416 min Scan# 872
Delta R.T. 0.018 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion: 117 Resp: 47155

Ion	Ratio	Lower	Upper
117	100		
119	571.0	77.4	116.2#
201	0.0	84.6	127.0#









#19
n-Nitroso-di-n-propylamine
Concen: 3.236 ng
RT: 7.304 min Scan# 81
Delta R.T. 0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:	70	Resp:	24024
Ion	Ratio	Lower	Upper
70	100		
42	35.5	52.3	78.5#
101	0.0	7.5	11.3#
130	0.0	16.1	24.1#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1000
Delta R.T. 0.000 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion:	136	Resp:	588197
Ion	Ratio	Lower	Upper
136	100		
137	11.7	8.8	13.2
54	11.1	7.5	11.3
68	9.7	4.7	7.1#









#22
Acetophenone
Concen: 2.567 ng
RT: 7.275 min Scan# 848
Delta R.T. -0.023 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:105 Resp: 35148
Ion Ratio Lower Upper
105 100
71 62.5 5.0 7.4#
51 20.8 25.0 37.4#
120 0.6 17.8 26.8#



#23
Nitrobenzene-d5
Concen: 89.015 ng
RT: 7.451 min Scan# 878
Delta R.T. 0.000 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion: 82 Resp: 959140
Ion Ratio Lower Upper
82 100
128 44.5 34.0 51.0
54 51.9 46.9 70.3









#25
Isophorone
Concen: 3.813 ng
RT: 7.687 min Scan# 918
Delta R.T. -0.023 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion: 82 Resp: 73650
Ion Ratio Lower Upper
82 100
95 174.1 5.7 8.5#
138 9.2 14.2 21.2#



#31
Naphthalene
Concen: 9.642 ng
RT: 8.192 min Scan# 1004
Delta R.T. 0.000 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion: 128 Resp: 288141
Ion Ratio Lower Upper
128 100
129 15.3 9.2 13.8#
127 15.5 10.5 15.7









#35
 Caprolactam
 Concen: 5.351 ng
 RT: 8.616 min Scan# 1076
 Delta R.T. -0.006 min
 Lab File: BF129458.D
 Acq: 30 Jul 2022 00:02




Instrument :
 BNA_F
 ClientSampleId :
 MHA


Tgt Ion:113 Resp: 14743
 Ion Ratio Lower Upper
 113 100
 55 1008.5 188.5 228.5#
 56 399.1 130.6 170.6#



#37
 2-Methylnaphthalene
 Concen: 56.375 ng
 RT: 8.887 min Scan# 1122
 Delta R.T. 0.006 min
 Lab File: BF129458.D
 Acq: 30 Jul 2022 00:02

Tgt Ion:142 Resp: 1094813
 Ion Ratio Lower Upper
 142 100
 141 88.7 71.0 106.4









#38
1-Methylnaphthalene
Concen: 38.372 ng
RT: 8.987 min Scan# 1138
Delta R.T. 0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:142 Resp: 719374
Ion Ratio Lower Upper
142 100
141 93.2 74.2 111.4



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.934 min Scan# 1300
Delta R.T. 0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion:164 Resp: 290866
Ion Ratio Lower Upper
164 100
162 102.1 79.5 119.3
160 46.6 34.2 51.4









#42
2,4,6-Tribromophenol
Concen: 111.822 ng
RT: 10.728 min Scan# 1434
Delta R.T. 0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:330 Resp: 312707
Ion Ratio Lower Upper
330 100
332 103.2 76.8 115.2
141 35.3 28.0 42.0



#45
2-Fluorobiphenyl
Concen: 80.515 ng
RT: 9.251 min Scan# 1184
Delta R.T. 0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion:172 Resp: 1513892
Ion Ratio Lower Upper
172 100
171 35.3 28.8 43.2
170 23.2 19.8 29.8









#46
1,1'-Biphenyl
Concen: 2.758 ng
RT: 9.345 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:154 Resp: 58542
Ion Ratio Lower Upper
154 100
153 43.6 20.6 60.6
76 14.5 0.0 33.3



#52
Acenaphthene
Concen: 3.823 ng
RT: 9.963 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion:154 Resp: 65110
Ion Ratio Lower Upper
154 100
153 128.4 89.6 134.4
152 81.6 42.4 63.6#









#53
3-Nitroaniline
Concen: 2.678 ng
RT: 9.881 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:138 Resp: 14213
Ion Ratio Lower Upper
138 100
108 52.6 10.9 16.3#
92 40.3 98.2 147.2#



#55
Dibenzofuran
Concen: 3.037 ng
RT: 10.139 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion:168 Resp: 71303
Ion Ratio Lower Upper
168 100
139 74.9 30.6 46.0#
169 120.1 11.0 16.6#









#56
4-Nitrophenol
Concen: 14.755 ng
RT: 10.034 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:139 Resp: 57769
Ion Ratio Lower Upper
139 100
109 104.7 50.7 90.7#
65 41.4 74.5 114.5#



#57
2,4-Dinitrotoluene
Concen: 5.468 ng
RT: 10.134 min Scan# 1334
Delta R.T. 0.012 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion:165 Resp: 32100
Ion Ratio Lower Upper
165 100
63 80.6 39.8 59.6#
89 38.5 60.2 90.2#
182 16.0 2.2 3.2#









#58
 Fluorene
 Concen: 7.369 ng
 RT: 10.481 min Scan# 1393
 Delta R.T. 0.000 min
 Lab File: BF129458.D
 Acq: 30 Jul 2022 00:02




Instrument :
 BNA_F
 ClientSampleId :
 MHA


Tgt Ion:166 Resp: 138429
 Ion Ratio Lower Upper
 166 100
 165 109.6 78.1 117.1
 167 22.0 10.5 15.7#



#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.422 min Scan# 1553
 Delta R.T. 0.006 min
 Lab File: BF129458.D
 Acq: 30 Jul 2022 00:02

Tgt Ion:188 Resp: 426122
 Ion Ratio Lower Upper
 188 100
 94 11.7 7.0 10.6#
 80 12.7 7.6 11.4#









#66
n-Nitrosodiphenylamine
Concen: 13.609 ng
RT: 10.592 min Scan# 1412
Delta R.T. 0.000 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:169 Resp: 193128
Ion Ratio Lower Upper
169 100
168 21.1 50.1 75.1#
167 15.3 27.9 41.9#



#71
Phenanthrene
Concen: 19.277 ng
RT: 11.445 min Scan# 1557
Delta R.T. 0.000 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion:178 Resp: 448405
Ion Ratio Lower Upper
178 100
176 19.6 16.0 24.0
179 18.5 12.6 19.0





#72

Anthracene

Concen: 19.616 ng

RT: 11.445 min Scan# 1557

Delta R.T. -0.053 min

Lab File: BF129458.D

Acq: 30 Jul 2022 00:02

Instrument :

BNA_F

ClientSampleId :

MHA





Tgt Ion:178 Resp: 448405

Ion Ratio Lower Upper

178 100

176 19.6 15.4 23.2

179 18.5 12.6 19.0



#75

Fluoranthene

Concen: 10.319 ng

RT: 12.633 min Scan# 1759

Delta R.T. 0.000 min

Lab File: BF129458.D

Acq: 30 Jul 2022 00:02




Tgt Ion:202 Resp: 243194


Ion Ratio Lower Upper

202 100

101 12.4 0.0 31.4

203 18.8 0.0 38.0









#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.063 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:240 Resp: 312400
Ion Ratio Lower Upper
240 100
120 10.6 8.9 13.3
236 25.0 20.8 31.2



#78
Pyrene
Concen: 8.660 ng
RT: 12.863 min Scan# 1798
Delta R.T. -0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion:202 Resp: 226242
Ion Ratio Lower Upper
202 100
200 21.4 17.7 26.5
203 19.0 14.6 21.8









#79
 Terphenyl-d14
 Concen: 71.601 ng
 RT: 13.004 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BF129458.D
 Acq: 30 Jul 2022 00:02




Instrument :
 BNA_F
 ClientSampleId :
 MHA


Tgt Ion:244 Resp: 1185649
 Ion Ratio Lower Upper
 244 100
 212 6.8 6.1 9.1
 122 11.7 8.8 13.2



#81
 Benzo(a)anthracene
 Concen: 5.641 ng
 RT: 14.051 min Scan# 2000
 Delta R.T. -0.006 min
 Lab File: BF129458.D
 Acq: 30 Jul 2022 00:02

Tgt Ion:228 Resp: 120384
 Ion Ratio Lower Upper
 228 100
 226 26.4 22.1 33.1
 229 18.9 15.8 23.8





#83

Chrysene

Concen: 4.757 ng

RT: 14.086 min Scan# 2006

Delta R.T. -0.006 min

Lab File: BF129458.D

Acq: 30 Jul 2022 00:02

Instrument :

BNA_F

ClientSampleId :

MHA





Tgt Ion:228 Resp: 95664

Ion Ratio Lower Upper

228 100

226 28.8 24.3 36.5

229 23.4 15.9 23.9



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.557 min Scan# 2256

Delta R.T. 0.006 min

Lab File: BF129458.D

Acq: 30 Jul 2022 00:02




Tgt Ion:264 Resp: 337019


Ion Ratio Lower Upper

264 100

260 23.8 18.4 27.6

265 21.7 17.2 25.8









#87
Indeno(1,2,3-cd)pyrene
Concen: 2.143 ng
RT: 17.062 min Scan# 2180
Delta R.T. -0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02




Instrument :
BNA_F
ClientSampleId :
MHA


Tgt Ion:276 Resp: 48645
Ion Ratio Lower Upper
276 100
138 22.0 21.1 31.7
277 24.4 20.3 30.5



#88
Benzo(b)fluoranthene
Concen: 8.319 ng
RT: 15.110 min Scan# 2180
Delta R.T. -0.006 min
Lab File: BF129458.D
Acq: 30 Jul 2022 00:02

Tgt Ion:252 Resp: 185981
Ion Ratio Lower Upper
252 100
253 21.5 17.7 26.5
125 13.0 8.0 12.0#





#89

Benzo(k)fluoranthene

Concen: 8.489 ng

RT: 15.110 min Scan# 2186

Delta R.T. -0.035 min

Lab File: BF129458.D

Acq: 30 Jul 2022 00:02

Instrument :

BNA_F

ClientSampleId :

MHA





Tgt Ion:252 Resp: 185981

Ion Ratio Lower Upper

252 100

253 21.5 17.8 26.6

125 13.0 8.3 12.5#



#90

Benzo(a)pyrene

Concen: 5.431 ng

RT: 15.486 min Scan# 2244

Delta R.T. -0.006 min

Lab File: BF129458.D

Acq: 30 Jul 2022 00:02




Tgt Ion:252 Resp: 98566


Ion Ratio Lower Upper

252 100

253 22.9 17.0 25.6

125 13.4 8.8 13.2#





#92

Benzo(g,h,i)perylene

Concen: 2.269 ng

RT: 17.521 min Scan# 2590

Delta R.T. -0.012 min

Lab File: BF129458.D


Acq: 30 Jul 2022 00:02

Instrument :

BNA_F

ClientSampleId :

MHA



Tgt Ion: 276 Resp: 42830

Ion Ratio Lower Upper

276 100

277 24.4 19.1 28.7

138 23.9 18.0 27.0

