

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072922\
 Data File : BF129436.D
 Acq On : 29 Jul 2022 11:52
 Operator : CG\JU
 Sample : N3895-13
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RVE-51

Quant Time: Jul 29 18:19:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	172553	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	696262	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	371720	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	617196	20.000	ng	# 0.00
76) Chrysene-d12	14.063	240	386301	20.000	ng	0.00
86) Perylene-d12	15.562	264	378439	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1185526	111.920	ng	0.02
7) Phenol-d6	6.522	99	1542063	115.821	ng	0.00
23) Nitrobenzene-d5	7.451	82	974254	76.384	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	425192	118.974	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1720754	71.611	ng	0.00
79) Terphenyl-d14	13.004	244	1622566	79.241	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072922\
Data File : BF129436.D
Acq On : 29 Jul 2022 11:52
Operator : CG\JU
Sample : N3895-13
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RVE-51

Quant Time: Jul 29 18:19:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 17:24:17 2022
Response via : Initial Calibration

