

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072922\
 Data File : BF129432.D
 Acq On : 29 Jul 2022 09:41
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 29 18:33:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	76	0.00
2	1,4-Dioxane	40.000	37.304	6.7	71	0.01
3	Pyridine	40.000	37.828	5.4	70	0.02
4	n-Nitrosodimethylamine	40.000	39.715	0.7	75	0.02
5 S	2-Fluorophenol	80.000	77.829	2.7	76	0.02
6	Aniline	40.000	37.653	5.9	72	0.00
7 S	Phenol-d6	80.000	78.951	1.3	77	0.01
8	2-Chlorophenol	40.000	39.473	1.3	76	0.00
9	Benzaldehyde	40.000	36.170	9.6	74	0.00
10 C	Phenol	40.000	40.879	-2.2	80	0.02
11	bis(2-Chloroethyl)ether	40.000	39.080	2.3	76	0.00
12	1,3-Dichlorobenzene	40.000	39.247	1.9	76	0.00
13 C	1,4-Dichlorobenzene	40.000	39.587	1.0	77	0.00
14	1,2-Dichlorobenzene	40.000	39.771	0.6	76	0.00
15	Benzyl Alcohol	40.000	41.173	-2.9	78	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.001	7.5	71	0.00
17	2-Methylphenol	40.000	38.706	3.2	75	0.01
18	Hexachloroethane	40.000	40.077	-0.2	77	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.626	3.4	76	0.00
20	3+4-Methylphenols	40.000	37.490	6.3	76	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	75	0.00
22	Acetophenone	40.000	39.341	1.6	78	0.00
23 S	Nitrobenzene-d5	80.000	81.309	-1.6	78	0.00
24	Nitrobenzene	40.000	39.915	0.2	75	0.00
25	Isophorone	40.000	39.001	2.5	76	0.00
26 C	2-Nitrophenol	40.000	40.310	-0.8	76	0.00
27	2,4-Dimethylphenol	40.000	39.887	0.3	77	0.01
28	bis(2-Chloroethoxy)methane	40.000	39.494	1.3	76	0.00
29 C	2,4-Dichlorophenol	40.000	40.287	-0.7	76	0.00
30	1,2,4-Trichlorobenzene	40.000	39.994	0.0	77	0.00
31	Naphthalene	40.000	40.179	-0.4	78	0.00
32	Benzoic acid	40.000	25.449	36.4#	46	0.00
33	4-Chloroaniline	40.000	39.542	1.1	77	0.00
34 C	Hexachlorobutadiene	40.000	41.948	-4.9	81	0.00
35	Caprolactam	40.000	39.564	1.1	76	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.936	-2.3	79	0.00
37	2-Methylnaphthalene	40.000	39.894	0.3	77	0.00
38	1-Methylnaphthalene	40.000	39.915	0.2	78	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	76	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.892	-2.2	80	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.757	5.6	67	0.00
42 S	2,4,6-Tribromophenol	80.000	80.825	-1.0	78	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.831	0.4	76	0.00
44	2,4,5-Trichlorophenol	40.000	39.530	1.2	77	0.01
45 S	2-Fluorobiphenyl	80.000	75.089	6.1	81	0.00
46	1,1'-Biphenyl	40.000	39.942	0.1	78	0.00
47	2-Chloronaphthalene	40.000	39.864	0.3	78	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072922\
 Data File : BF129432.D
 Acq On : 29 Jul 2022 09:41
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 29 18:33:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 17:24:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.543	-1.4	77	0.00
49	Acenaphthylene	40.000	40.024	-0.1	79	0.00
50	Dimethylphthalate	40.000	40.301	-0.8	79	0.00
51	2,6-Dinitrotoluene	40.000	40.368	-0.9	77	0.00
52 C	Acenaphthene	40.000	39.444	1.4	76	0.00
53	3-Nitroaniline	40.000	38.743	3.1	73	0.00
54 P	2,4-Dinitrophenol	40.000	36.890	7.8	67	0.00
55	Dibenzofuran	40.000	40.039	-0.1	78	0.00
56 P	4-Nitrophenol	40.000	34.079	14.8	64	0.02
57	2,4-Dinitrotoluene	40.000	40.999	-2.5	76	0.00
58	Fluorene	40.000	40.281	-0.7	80	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.776	3.1	73	0.01
60	Diethylphthalate	40.000	40.084	-0.2	78	0.00
61	4-Chlorophenyl-phenylether	40.000	40.237	-0.6	80	0.00
62	4-Nitroaniline	40.000	38.887	2.8	75	0.00
63	Azobenzene	40.000	39.045	2.4	76	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	76	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.888	0.3	72	0.01
66 c	n-Nitrosodiphenylamine	40.000	39.739	0.7	77	0.00
67	4-Bromophenyl-phenylether	40.000	40.230	-0.6	78	0.00
68	Hexachlorobenzene	40.000	40.756	-1.9	79	0.00
69	Atrazine	40.000	39.788	0.5	77	0.00
70 C	Pentachlorophenol	40.000	38.475	3.8	70	0.00
71	Phenanthrene	40.000	39.215	2.0	77	0.00
72	Anthracene	40.000	39.783	0.5	79	0.00
73	Carbazole	40.000	39.455	1.4	77	0.00
74	Di-n-butylphthalate	40.000	40.150	-0.4	79	0.00
75 C	Fluoranthene	40.000	40.077	-0.2	79	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	80	0.00
77	Benzydine	40.000	35.561	11.1	68	0.00
78	Pyrene	40.000	39.352	1.6	79	0.00
79 S	Terphenyl-d14	80.000	79.771	0.3	82	0.00
80	Butylbenzylphthalate	40.000	41.096	-2.7	80	0.00
81	Benzo(a)anthracene	40.000	39.575	1.1	80	0.00
82	3,3'-Dichlorobenzidine	40.000	38.570	3.6	77	0.00
83	Chrysene	40.000	39.285	1.8	81	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	40.062	-0.2	80	0.00
85 c	Di-n-octyl phthalate	40.000	41.738	-4.3	85	0.00
86 I	Perylene-d12	20.000	20.000	0.0	80	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	39.129	2.2	72	0.02
88	Benzo(b)fluoranthene	40.000	39.433	1.4	82	0.01
89	Benzo(k)fluoranthene	40.000	40.327	-0.8	79	0.01
90 C	Benzo(a)pyrene	40.000	39.564	1.1	79	0.01
91	Dibenzo(a,h)anthracene	40.000	39.394	1.5	72	0.02
92	Benzo(g,h,i)perylene	40.000	39.276	1.8	70	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072922\
Data File : BF129432.D
Acq On : 29 Jul 2022 09:41
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 29 18:33:43 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 17:24:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0