

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073021\
 Data File : BF124884.D
 Acq On : 30 Jul 2021 19:31
 Operator : JU/CG
 Sample : M3234-01 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OR-03-072921

Quant Time: Jul 31 00:07:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 15:28:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	6655	20.00	ng	0.00
21) Naphthalene-d8	8.151	136	27060	20.00	ng	# 0.00
39) Acenaphthene-d10	9.910	164	15025	20.00	ng	0.00
64) Phenanthrene-d10	11.404	188	23213	20.00	ng	0.00
76) Chrysene-d12	14.051	240	13633	20.00	ng	# 0.01
86) Perylene-d12	15.527	264	11443	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	7004	17.82	ng	0.00
7) Phenol-d6	6.493	99	9077	18.42	ng	0.00
23) Nitrobenzene-d5	7.428	82	6188	13.61	ng	-0.01
42) 2,4,6-Tribromophenol	10.698	330	2379	18.44	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	13507	13.20	ng	0.00
79) Terphenyl-d14	12.980	244	10450	13.14	ng	0.00
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	7.428	70	724	2.75	ng	# 79
25) Isophorone	7.428	82	6188	7.70	ng	# 59
31) Naphthalene	8.175	128	5257	3.72	ng	97
37) 2-Methylnaphthalene	8.863	142	3120	3.31	ng	94
38) 1-Methylnaphthalene	8.963	142	2816	3.15	ng	99
49) Acenaphthylene	9.769	152	7436	5.43	ng	98
51) 2,6-Dinitrotoluene	9.910	165	2018	8.83	ng	# 26
52) Acenaphthene	9.945	154	11444	12.61	ng	95
55) Dibenzofuran	10.116	168	17541	13.52	ng	98
56) 4-Nitrophenol	10.116	139	7131	37.55	ng	# 10
57) 2,4-Dinitrotoluene	9.910	165	2018	6.51	ng	# 18
58) Fluorene	10.463	166	25793	25.92	ng	97
71) Phenanthrene	11.439	178	206274	165.22	ng	98
72) Anthracene	11.480	178	57960	46.44	ng	98
73) Carbazole	11.627	167	9950	8.50	ng	# 96
75) Fluoranthene	12.633	202	195846	153.72	ng	97
78) Pyrene	12.857	202	153251	141.99	ng	97
81) Benzo(a)anthracene	14.039	228	73903	82.93	ng	99
83) Chrysene	14.074	228	56087	65.33	ng	97
87) Indeno(1,2,3-cd)pyrene	17.015	276	18932	28.66	ng	# 91
88) Benzo(b)fluoranthene	15.098	252	81761	101.48	ng	97
89) Benzo(k)fluoranthene	15.098	252	81761	111.07	ng	99
90) Benzo(a)pyrene	15.468	252	41149	61.14	ng	98
91) Dibenzo(a,h)anthracene	17.021	278	5507	9.52	ng	95
92) Benzo(g,h,i)perylene	17.468	276	15407	28.09	ng	# 85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073021\
Data File : BF124884.D
Acq On : 30 Jul 2021 19:31
Operator : JU/CG
Sample : M3234-01 5X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OR-03-072921

Quant Time: Jul 31 00:07:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 22 15:28:57 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.869 min Scan# 813
Delta R.T. -0.000 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921

Tgt Ion:152 Resp: 6655
Ion Ratio Lower Upper
152 100
150 148.7 117.6 176.4
115 60.5 47.1 70.7



#5
2-Fluorophenol
Concen: 17.82 ng
RT: 5.481 min Scan# 577
Delta R.T. 0.000 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:112 Resp: 7004
Ion Ratio Lower Upper
112 100
64 57.5 44.2 66.2
63 32.6 26.2 39.2









#7
Phenol-d6
Concen: 18.42 ng
RT: 6.493 min Scan# 749
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921

Tgt Ion:	99	Resp:	9077
Ion	Ratio	Lower	Upper
99	100		
42	22.6	23.5	35.3#
71	33.4	25.6	38.4



#19
n-Nitroso-di-n-propylamine
Concen: 2.75 ng
RT: 7.428 min Scan# 908
Delta R.T. 0.135 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:	70	Resp:	724
Ion	Ratio	Lower	Upper
70	100		
42	80.2	72.2	108.2
101	0.0	7.8	11.8#
130	0.0	21.8	32.6#








#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.151 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921

Tgt Ion:136	Resp:	27060
Ion Ratio	Lower	Upper
136	100	
137	10.6	9.8 14.6
54	9.0	6.4 9.6
68	5.3	3.3 4.9#



#23
Nitrobenzene-d5
Concen: 13.61 ng
RT: 7.428 min Scan# 908
Delta R.T. -0.012 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion: 82	Resp:	6188
Ion Ratio	Lower	Upper
82	100	
128	44.7	40.6 60.8
54	63.7	49.6 74.4









#25
Isophorone
Concen: 7.70 ng
RT: 7.428 min Scan# 908
Delta R.T. -0.271 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion: 82 Resp: 6188
Ion Ratio Lower Upper
82 100
95 0.0 6.5 9.7#
138 0.0 18.4 27.6#



#31
Naphthalene
Concen: 3.72 ng
RT: 8.175 min Scan# 1035
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion: 128 Resp: 5257
Ion Ratio Lower Upper
128 100
129 11.6 9.4 14.0
127 11.6 10.9 16.3









#37
2-Methylnaphthalene
Concen: 3.31 ng
RT: 8.863 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:142 Resp: 3120
Ion Ratio Lower Upper
142 100
141 83.6 71.2 106.8



#38
1-Methylnaphthalene
Concen: 3.15 ng
RT: 8.963 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:142 Resp: 2816
Ion Ratio Lower Upper
142 100
141 94.5 76.0 114.0









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.910 min Scan# 1330
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:164	Resp:	15025
Ion Ratio	Lower	Upper
164	100	
162	101.2	81.3 121.9
160	45.2	38.0 57.0



#42
2,4,6-Tribromophenol
Concen: 18.44 ng
RT: 10.698 min Scan# 1464
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:330	Resp:	2379
Ion Ratio	Lower	Upper
330	100	
332	79.3	76.2 114.4
141	39.7	44.6 66.8#









#45
2-Fluorobiphenyl
Concen: 13.20 ng
RT: 9.228 min Scan# 1214
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:172	Resp:	13507
Ion Ratio	Lower	Upper
172	100	
171	34.1	26.5 39.7
170	25.6	18.8 28.2



#49
Acenaphthylene
Concen: 5.43 ng
RT: 9.769 min Scan# 1306
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:152	Resp:	7436
Ion Ratio	Lower	Upper
152	100	
151	19.4	15.7 23.5
153	14.0	9.8 14.8









#51
2,6-Dinitrotoluene
Concen: 8.83 ng
RT: 9.910 min Scan# 1330
Delta R.T. 0.212 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampled :
OR-03-072921


Tgt Ion:165 Resp: 2018
Ion Ratio Lower Upper
165 100
63 0.0 42.6 63.8#
89 0.0 40.6 60.8#



#52
Acenaphthene
Concen: 12.61 ng
RT: 9.945 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:154 Resp: 11444
Ion Ratio Lower Upper
154 100
153 117.4 89.4 134.2
152 53.9 40.5 60.7









#55
Dibenzofuran
Concen: 13.52 ng
RT: 10.116 min Scan# 1365
Delta R.T. -0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:168 Resp: 17541
Ion Ratio Lower Upper
168 100
139 42.7 32.7 49.1
169 13.4 11.2 16.8



#56
4-Nitrophenol
Concen: 37.55 ng
RT: 10.116 min Scan# 1365
Delta R.T. 0.100 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:139 Resp: 7131
Ion Ratio Lower Upper
139 100
109 0.0 60.4 100.4#
65 0.0 58.4 98.4#









#57
2,4-Dinitrotoluene
Concen: 6.51 ng
RT: 9.910 min Scan# 1330
Delta R.T. -0.188 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:165	Resp:	2018
Ion Ratio	Lower	Upper
165	100	
63	0.0	40.9 61.3#
89	0.0	59.7 89.5#
182	0.0	0.0 0.0



#58
Fluorene
Concen: 25.92 ng
RT: 10.463 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:166	Resp:	25793
Ion Ratio	Lower	Upper
166	100	
165	99.9	78.1 117.1
167	15.7	11.1 16.7









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.404 min Scan# 1584
Delta R.T. 0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:188 Resp: 23213
Ion Ratio Lower Upper
188 100
94 7.2 6.7 10.1
80 9.5 7.5 11.3



#71
Phenanthrene
Concen: 165.22 ng
RT: 11.439 min Scan# 1590
Delta R.T. 0.012 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:178 Resp: 206274
Ion Ratio Lower Upper
178 100
176 19.3 14.7 22.1
179 15.4 11.5 17.3









#72
 Anthracene
 Concen: 46.44 ng
 RT: 11.480 min Scan# 1597
 Delta R.T. 0.006 min
 Lab File: BF124884.D
 Acq: 30 Jul 2021 19:31




Instrument :
 BNA_F
 ClientSampleId :
 OR-03-072921


Tgt Ion:178 Resp: 57960
 Ion Ratio Lower Upper
 178 100
 176 18.9 14.3 21.5
 179 15.1 12.2 18.4



#73
 Carbazole
 Concen: 8.50 ng
 RT: 11.627 min Scan# 1622
 Delta R.T. 0.000 min
 Lab File: BF124884.D
 Acq: 30 Jul 2021 19:31

Tgt Ion:167 Resp: 9950
 Ion Ratio Lower Upper
 167 100
 166 22.2 17.4 26.0
 139 16.9 11.0 16.4#









#75
Fluoranthene
Concen: 153.72 ng
RT: 12.633 min Scan# 1793
Delta R.T. 0.024 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:202 Resp: 195846
Ion Ratio Lower Upper
202 100
101 7.7 0.0 30.7
203 16.3 0.0 36.3



#76
Chrysene-d12
Concen: 20.00 ng
RT: 14.051 min Scan# 2034
Delta R.T. 0.012 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:240 Resp: 13633
Ion Ratio Lower Upper
240 100
120 10.6 10.8 16.2#
236 27.7 21.6 32.4





#78

Pyrene

Concen: 141.99 ng

RT: 12.857 min Scan# 1831

Delta R.T. 0.018 min

Lab File: BF124884.D

Acq: 30 Jul 2021 19:31

Tgt Ion:202 Resp: 153251





Ion	Ratio	Lower	Upper
202	100		
200	20.1	15.3	22.9
203	18.0	13.4	20.0

Instrument :

BNA_F

ClientSampleId :

OR-03-072921



#79

Terphenyl-d14

Concen: 13.14 ng

RT: 12.980 min Scan# 1852




Delta R.T. -0.006 min


Lab File: BF124884.D

Acq: 30 Jul 2021 19:31

Tgt Ion:244 Resp: 10450

Ion	Ratio	Lower	Upper
244	100		
212	7.6	5.9	8.9
122	10.2	9.0	13.4









#81
Benzo(a)anthracene
Concen: 82.93 ng
RT: 14.039 min Scan# 2032
Delta R.T. 0.012 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:228 Resp: 73903
Ion Ratio Lower Upper
228 100
226 28.0 21.6 32.4
229 19.8 15.6 23.4



#83
Chrysene
Concen: 65.33 ng
RT: 14.074 min Scan# 2038
Delta R.T. 0.006 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:228 Resp: 56087
Ion Ratio Lower Upper
228 100
226 29.0 22.6 34.0
229 22.0 15.9 23.9









#86
Perylene-d12
Concen: 20.00 ng
RT: 15.527 min Scan# 2285
Delta R.T. 0.018 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:264	Resp:	11443
Ion Ratio	Lower	Upper
264	100	
260	21.2	19.6 29.4
265	23.9	17.2 25.8



#87
Indeno(1,2,3-cd)pyrene
Concen: 28.66 ng
RT: 17.015 min Scan# 2538
Delta R.T. 0.029 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:276	Resp:	18932
Ion Ratio	Lower	Upper
276	100	
138	21.7	24.1 36.1#
277	23.9	18.8 28.2









#88
Benzo(b)fluoranthene
Concen: 101.48 ng
RT: 15.098 min Scan# 2212
Delta R.T. 0.018 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:252 Resp: 81761
Ion Ratio Lower Upper
252 100
253 22.1 16.8 25.2
125 10.6 9.8 14.8



#89
Benzo(k)fluoranthene
Concen: 111.07 ng
RT: 15.098 min Scan# 2212
Delta R.T. -0.012 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Tgt Ion:252 Resp: 81761
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 10.6 7.2 10.8









#90
Benzo(a)pyrene
Concen: 61.14 ng
RT: 15.468 min Scan# 2275
Delta R.T. 0.024 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31




Instrument :
BNA_F
ClientSampleId :
OR-03-072921


Tgt Ion:	252	Resp:	41149
Ion Ratio	Lower	Upper	
252	100		
253	22.1	18.4	27.6
125	11.0	9.9	14.9



#91
Dibenzo(a,h)anthracene
Concen: 9.52 ng
RT: 17.021 min Scan# 2539
Delta R.T. 0.018 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31


Tgt Ion:	278	Resp:	5507
Ion Ratio	Lower	Upper	
278	100		
139	18.5	14.9	22.3
279	28.7	19.7	29.5





#92
Benzo(g,h,i)perylene
Concen: 28.09 ng
RT: 17.468 min Scan# 2615
Delta R.T. 0.035 min
Lab File: BF124884.D
Acq: 30 Jul 2021 19:31

Instrument :
BNA_F
ClientSampleId :
OR-03-072921



Tgt Ion	Ratio	Lower	Upper
276	100		
277	21.9	22.0	33.0#
138	18.0	22.2	33.4#

