




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073024\
Data File : BF138679.D
Acq On : 30 Jul 2024 12:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 30 18:29:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration



TIC: BF138679.D\data.ms

(70) Pentachlorophenol (C)




11.174min (-0.001) 37845.67 ng

response 69604

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.50	63.56
264.00	62.20	62.86
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073024\
Data File : BF138679.D
Acq On : 30 Jul 2024 12:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 30 18:29:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration



TIC: BF138679.D\data.ms

(77) Benzidine

12.704min (-0.000) 0.00 ng

response 376863

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	13.91
183.00	12.00	12.39
0.00	0.00	0.00