

Data Path : Z:\HPCHEM1\BNA F\DATA\BF073116\
 Data File : BF089466.D
 Acq On : 31 Jul 2016 12:52
 Operator : UM/SJ
 Sample : H4284-17
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AB-SLOPE(0-4)I

Quant Time: Aug 01 05:13:59 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 01 04:11:30 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.49	152	47509	20.00	ng	-0.01
21) Naphthalene-d8	7.78	136	182333	20.00	ng	-0.01
38) Acenaphthene-d10	9.52	164	84438	20.00	ng	-0.01
63) Phenanthrene-d10	10.99	188	160908	20.00	ng	-0.01
75) Chrysene-d12	13.62	240	93835	20.00	ng	0.00
86) Perylene-d12	14.94	264	81830	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.07	112	388663	130.63	ng	0.01
7) Phenol-d6	6.17	99	522785	132.95	ng	0.00
23) Nitrobenzene-d5	7.07	82	332336	107.59	ng	-0.01
41) 2,4,6-Tribromophenol	10.31	330	75967	108.63	ng	-0.01
44) 2-Fluorobiphenyl	8.86	172	466059	81.00	ng	-0.01
78) Terphenyl-d14	12.58	244	332374	82.90	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.27	163	72204	11.24	ng	99
73) Di-n-butylphthalate	11.56	149	43553	4.18	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF073116\
Data File : BF089466.D
Acq On : 31 Jul 2016 12:52
Operator : UM/SJ
Sample : H4284-17
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
AB-SLOPE(0-4)I

Quant Time: Aug 01 05:13:59 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF072716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 01 04:11:30 2016
Response via : Initial Calibration

