


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073120\
 Data File : BF121147.D
 Acq On : 31 Jul 2020 9:29
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF071620.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 29 14:55:43 2020



AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1608

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.1	31925	PASS
68	69	0.00	2	1.8	561	PASS
69	198	0.00	100	33.4	30432	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.1	43842	PASS
197	198	0.00	2	0.5	487	PASS
198	198	100	100	100.0	91064	PASS
199	198	5	9	7.1	6433	PASS
275	198	10	60	27.6	25092	PASS
365	198	1	100	3.5	3146	PASS
441	198	0.01	100	12.3	11230	PASS
442	198	50	100	77.8	70853	PASS
443	442	15	24	18.6	13172	PASS