

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073124\
 Data File : BF138691.D
 Acq On : 31 Jul 2024 09:00
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 31 09:24:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:50:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	41.357	-3.4	103	0.01
3	Pyridine	40.000	40.884	-2.2	103	0.00
4	n-Nitrosodimethylamine	40.000	39.553	1.1	101	0.01
5 S	2-Fluorophenol	80.000	79.964	0.0	103	0.00
6	Aniline	40.000	41.224	-3.1	106	0.00
7 S	Phenol-d6	80.000	78.808	1.5	103	0.00
8	2-Chlorophenol	40.000	40.108	-0.3	104	0.00
9	Benzaldehyde	40.000	32.996	17.5	96	0.00
10 C	Phenol	40.000	39.676	0.8	104	0.00
11	bis(2-Chloroethyl)ether	40.000	38.836	2.9	102	0.00
12	1,3-Dichlorobenzene	40.000	39.465	1.3	103	0.00
13 C	1,4-Dichlorobenzene	40.000	39.355	1.6	103	0.00
14	1,2-Dichlorobenzene	40.000	39.729	0.7	103	0.00
15	Benzyl Alcohol	40.000	40.339	-0.8	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.936	2.7	102	0.00
17	2-Methylphenol	40.000	39.460	1.3	102	0.00
18	Hexachloroethane	40.000	38.953	2.6	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.961	2.6	105	0.00
20	3+4-Methylphenols	40.000	38.724	3.2	104	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	39.265	1.8	105	0.00
23 S	Nitrobenzene-d5	80.000	78.724	1.6	104	0.00
24	Nitrobenzene	40.000	39.177	2.1	103	0.00
25	Isophorone	40.000	39.518	1.2	107	0.00
26 C	2-Nitrophenol	40.000	40.898	-2.2	105	0.00
27	2,4-Dimethylphenol	40.000	39.820	0.4	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.599	1.0	106	0.00
29 C	2,4-Dichlorophenol	40.000	40.190	-0.5	106	0.00
30	1,2,4-Trichlorobenzene	40.000	39.329	1.7	104	0.00
31	Naphthalene	40.000	39.014	2.5	104	0.00
32	Benzoic acid	40.000	39.839	0.4	107	0.00
33	4-Chloroaniline	40.000	40.190	-0.5	108	0.00
34 C	Hexachlorobutadiene	40.000	39.691	0.8	106	0.00
35	Caprolactam	40.000	41.065	-2.7	114	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.898	0.3	108	0.00
37	2-Methylnaphthalene	40.000	39.080	2.3	106	0.00
38	1-Methylnaphthalene	40.000	39.131	2.2	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	111	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.947	2.6	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.690	3.3	112	0.00
42 S	2,4,6-Tribromophenol	80.000	79.030	1.2	113	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.702	3.2	108	0.00
44	2,4,5-Trichlorophenol	40.000	39.935	0.2	111	0.00
45 S	2-Fluorobiphenyl	80.000	75.657	5.4	107	0.00
46	1,1'-Biphenyl	40.000	38.177	4.6	108	0.00
47	2-Chloronaphthalene	40.000	38.116	4.7	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073124\
 Data File : BF138691.D
 Acq On : 31 Jul 2024 09:00
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 31 09:24:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:50:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.505	1.2	111	0.00
49	Acenaphthylene	40.000	39.327	1.7	110	0.00
50	Dimethylphthalate	40.000	39.454	1.4	114	0.00
51	2,6-Dinitrotoluene	40.000	39.823	0.4	111	0.00
52 C	Acenaphthene	40.000	38.357	4.1	109	0.00
53	3-Nitroaniline	40.000	39.525	1.2	113	0.00
54 P	2,4-Dinitrophenol	40.000	38.574	3.6	112	0.00
55	Dibenzofuran	40.000	38.723	3.2	111	0.00
56 P	4-Nitrophenol	40.000	41.015	-2.5	116	0.00
57	2,4-Dinitrotoluene	40.000	40.335	-0.8	114	0.00
58	Fluorene	40.000	38.991	2.5	112	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.667	0.8	114	0.00
60	Diethylphthalate	40.000	40.205	-0.5	117	0.00
61	4-Chlorophenyl-phenylether	40.000	38.591	3.5	111	0.00
62	4-Nitroaniline	40.000	39.735	0.7	115	0.00
63	Azobenzene	40.000	39.511	1.2	113	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	116	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.855	-2.1	116	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.112	-0.3	115	0.00
67	4-Bromophenyl-phenylether	40.000	39.397	1.5	114	0.00
68	Hexachlorobenzene	40.000	39.625	0.9	117	0.00
69	Atrazine	40.000	40.834	-2.1	118	0.00
70 C	Pentachlorophenol	40.000	41.835	-4.6	120	0.00
71	Phenanthrene	40.000	39.340	1.6	115	0.00
72	Anthracene	40.000	39.563	1.1	116	0.00
73	Carbazole	40.000	38.789	3.0	114	0.00
74	Di-n-butylphthalate	40.000	41.037	-2.6	118	0.00
75 C	Fluoranthene	40.000	39.151	2.1	113	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	42.353	-5.9	95	0.00
78	Pyrene	40.000	40.840	-2.1	110	0.00
79 S	Terphenyl-d14	80.000	81.952	-2.4	111	0.00
80	Butylbenzylphthalate	40.000	41.760	-4.4	101	0.00
81	Benzo(a)anthracene	40.000	39.431	1.4	98	0.00
82	3,3'-Dichlorobenzidine	40.000	39.940	0.2	98	0.00
83	Chrysene	40.000	40.941	-2.4	105	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.990	-2.5	96	0.00
85 c	Di-n-octyl phthalate	40.000	40.783	-2.0	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.454	1.4	100	0.00
88	Benzo(b)fluoranthene	40.000	36.603	8.5	91	0.00
89	Benzo(k)fluoranthene	40.000	39.693	0.8	105	0.00
90 C	Benzo(a)pyrene	40.000	39.159	2.1	99	0.00
91	Dibenzo(a,h)anthracene	40.000	39.251	1.9	101	0.00
92	Benzo(g,h,i)perylene	40.000	38.803	3.0	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073124\
Data File : BF138691.D
Acq On : 31 Jul 2024 09:00
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 31 09:24:56 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0