



Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF073124\  
Data File : BF138690.D  
Acq On : 31 Jul 2024 08:31  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF073024.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Tue Jul 30 17:50:01 2024






AutoFind: Scans 1618, 1619, 1620; Background Corrected with Scan 1612

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 48.9         | 44816      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.8          | 671        | PASS                |
| 69             | 198             | 0.00            | 100             | 39.6         | 36293      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.6          | 227        | PASS                |
| 127            | 198             | 10              | 80              | 48.6         | 44493      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 91587      | PASS                |
| 199            | 198             | 5               | 9               | 6.7          | 6138       | PASS                |
| 275            | 198             | 10              | 60              | 24.7         | 22640      | PASS                |
| 365            | 198             | 1               | 100             | 2.6          | 2345       | PASS                |
| 441            | 198             | 0.01            | 100             | 11.9         | 10904      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 67619      | PASS                |
| 443            | 442             | 15              | 24              | 18.6         | 12566      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF073124\  
Data File : BF138690.D  
Acq On : 31 Jul 2024 08:31  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Jul 31 11:03:11 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF073024.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Jul 30 17:50:01 2024  
Response via : Initial Calibration



TIC: BF138690.D\data.ms

**(70) Pentachlorophenol (C)**

**11.175min (-0.000) 281854.08 ng**




**response 78111**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 61.50  | 62.97  |
| 264.00 | 62.20  | 60.56  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF073124\  
Data File : BF138690.D  
Acq On : 31 Jul 2024 08:31  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Jul 31 11:03:11 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF073024.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Jul 30 17:50:01 2024  
Response via : Initial Calibration



TIC: BF138690.D\data.ms

(77) Benzidine

12.704min (-0.000) 80204.41 ng

response 454649

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 13.90  | 14.18  |
| 183.00 | 12.00  | 12.30  |
| 0.00   | 0.00   | 0.00   |

### DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 7/31/2024     | BNA_F            | <u>BF138690.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 232281           | 13.521             |
| DDD           | 5521             | 13.227             |
| DDE           | 195              | 12.892             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 5716          | 237997           | 2.40               |
|               |                  |                    |

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP