

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080119\
 Data File : BF115900.D
 Acq On : 1 Aug 2019 11:52
 Operator : HP/JU
 Sample : K3621-07 5X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OK-01-072919

Quant Time: Aug 01 14:02:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 15:31:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	142442	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	559761	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	285084	20.00	ng	0.00
64) Phenanthrene-d10	11.38	188	425081	20.00	ng	0.00
76) Chrysene-d12	14.02	240	323862	20.00	ng	0.00
87) Perylene-d12	15.50	264	338080	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	180443	21.21	ng	0.00
7) Phenol-d6	6.48	99	244414	22.77	ng	0.00
23) Nitrobenzene-d5	7.41	82	146815	15.14	ng	0.00
42) 2,4,6-Tribromophenol	10.68	330	53862	19.49	ng	0.00
45) 2-Fluorobiphenyl	9.21	172	301519	19.81	ng	0.00
79) Terphenyl-d14	12.96	244	274960	17.89	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.60	163	73407	3.782	ng	98
75) Fluoranthene	12.60	202	50432	2.217	ng	97
78) Pyrene	12.83	202	52678	2.158	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080119\
Data File : BF115900.D
Acq On : 1 Aug 2019 11:52
Operator : HP/JU
Sample : K3621-07 5X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OK-01-072919

Quant Time: Aug 01 14:02:50 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 15:31:51 2019
Response via : Initial Calibration

