

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080119\
 Data File : BF115925.D
 Acq On : 1 Aug 2019 23:42
 Operator : HP/JU
 Sample : K4044-17
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-3

Manual Integrations
 APPROVED

mohammad
 8/5/2019 1:48:17 PM

Quant Time: Aug 02 07:26:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 15:31:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	143343	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	536931	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	272237	20.00	ng	0.00
64) Phenanthrene-d10	11.38	188	409575	20.00	ng	0.00
76) Chrysene-d12	14.02	240	296294	20.00	ng	0.00
87) Perylene-d12	15.49	264	327355	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	1078552	125.96	ng	0.01
7) Phenol-d6	6.49	99	1393906	129.03	ng	0.00
23) Nitrobenzene-d5	7.42	82	885654	95.21	ng	0.00
42) 2,4,6-Tribromophenol	10.69	330	321000	121.62	ng	0.00
45) 2-Fluorobiphenyl	9.22	172	1363138	93.79	ng	0.00
79) Terphenyl-d14	12.96	244	1155699	82.19	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.60	163	214488	11.571	ng	98
75) Fluoranthene	12.59	202	47812	2.181	ng	99
78) Pyrene	12.82	202	47740	2.137	ng	96
88) Benzo(b)fluoranthene	15.06	252	41493m	2.192	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080119\
Data File : BF115925.D
Acq On : 1 Aug 2019 23:42
Operator : HP/JU
Sample : K4044-17
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-3

Manual Integrations
APPROVED

mohammad
8/5/2019 1:48:17 PM

Quant Time: Aug 02 07:26:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 15:31:51 2019
Response via : Initial Calibration

