

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080123\
 Data File : BF134545.D
 Acq On : 01 Aug 2023 13:23
 Operator : CG\JU
 Sample : PB154403BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB154403BL

Quant Time: Aug 02 02:42:52 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 02:20:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.792	152	193009	20.000	ng	0.00
21) Naphthalene-d8	8.069	136	773216	20.000	ng	# 0.00
39) Acenaphthene-d10	9.822	164	400727	20.000	ng	0.00
64) Phenanthrene-d10	11.310	188	752713	20.000	ng	# 0.00
76) Chrysene-d12	13.951	240	592221	20.000	ng	#-0.01
86) Perylene-d12	15.409	264	543668	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	1607672	126.617	ng	0.02
7) Phenol-d6	6.422	99	2055763	126.658	ng	0.00
23) Nitrobenzene-d5	7.351	82	1239363	100.168	ng	0.00
42) 2,4,6-Tribromophenol	10.616	330	551423	134.986	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	2160598	89.647	ng	0.00
79) Terphenyl-d14	12.910	244	2608718	77.654	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080123\
Data File : BF134545.D
Acq On : 01 Aug 2023 13:23
Operator : CG\JU
Sample : PB154403BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB154403BL

Quant Time: Aug 02 02:42:52 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 02:20:34 2023
Response via : Initial Calibration

