

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080221\
 Data File : BF124898.D
 Acq On : 2 Aug 2021 9:31
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 11:23:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 11:15:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.413	0.468	-13.3	106	0.00
3	Pyridine	0.977	1.097	-12.3	103	0.00
4	n-Nitrosodimethylamine	0.773	0.767	0.8	94	0.00
5 S	2-Fluorophenol	1.181	1.177	0.3	98	0.00
6	Aniline	1.522	1.581	-3.9	103	0.00
7 S	Phenol-d6	1.481	1.515	-2.3	100	0.00
8	2-Chlorophenol	1.332	1.301	2.3	98	0.00
9	Benzaldehyde	0.797	0.776	2.6	103	0.00
10 C	Phenol	1.418	1.566	-10.4	112	0.00
11	bis(2-Chloroethyl)ether	1.124	1.189	-5.8	107	0.00
12	1,3-Dichlorobenzene	1.448	1.393	3.8	95	0.00
13 C	1,4-Dichlorobenzene	1.449	1.423	1.8	97	0.00
14	1,2-Dichlorobenzene	1.400	1.343	4.1	95	0.00
15	Benzyl Alcohol	0.974	1.115	-14.5	111	0.00
16	2,2'-oxybis(1-Chloropropane	1.892	1.894	-0.1	99	0.00
17	2-Methylphenol	0.971	0.954	1.8	96	0.00
18	Hexachloroethane	0.549	0.570	-3.8	101	0.00
19 P	n-Nitroso-di-n-propylamine	0.791	0.871	-10.1	106	0.00
20	3+4-Methylphenols	1.282	1.288	-0.5	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
22	Acetophenone	0.463	0.495	-6.9	100	0.00
23 S	Nitrobenzene-d5	0.336	0.385	-14.6	107	0.00
24	Nitrobenzene	0.329	0.382	-16.1	110	0.00
25	Isophorone	0.594	0.664	-11.8	105	0.00
26 C	2-Nitrophenol	0.182	0.194	-6.6	95	0.00
27	2,4-Dimethylphenol	0.281	0.266	5.3	89	0.00
28	bis(2-Chloroethoxy)methane	0.345	0.380	-10.1	105	0.00
29 C	2,4-Dichlorophenol	0.297	0.288	3.0	91	0.00
30	1,2,4-Trichlorobenzene	0.326	0.336	-3.1	96	0.00
31	Naphthalene	1.044	1.038	0.6	94	0.00
32	Benzoic acid	0.217	0.144	33.6#	60	0.00
33	4-Chloroaniline	0.392	0.392	0.0	95	0.00
34 C	Hexachlorobutadiene	0.195	0.208	-6.7	101	0.00
35	Caprolactam	0.077	0.078	-1.3	97	0.00
36 C	4-Chloro-3-methylphenol	0.284	0.319	-12.3	99	0.00
37	2-Methylnaphthalene	0.697	0.676	3.0	91	0.00
38	1-Methylnaphthalene	0.661	0.652	1.4	93	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	0.558	0.567	-1.6	99	0.00
41 P	Hexachlorocyclopentadiene	0.330	0.271	17.9	73	0.00
42 S	2,4,6-Tribromophenol	0.172	0.177	-2.9	97	0.00
43 C	2,4,6-Trichlorophenol	0.396	0.378	4.5	91	0.00
44	2,4,5-Trichlorophenol	0.407	0.396	2.7	90	0.00
45 S	2-Fluorobiphenyl	1.362	1.337	1.8	95	0.00
46	1,1'-Biphenyl	1.493	1.459	2.3	95	0.00
47	2-Chloronaphthalene	1.182	1.191	-0.8	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080221\
 Data File : BF124898.D
 Acq On : 2 Aug 2021 9:31
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 11:23:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 11:15:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.309	0.346	-12.0	105	0.00
49	Acenaphthylene	1.823	1.783	2.2	94	0.00
50	Dimethylphthalate	1.377	1.345	2.3	92	0.00
51	2,6-Dinitrotoluene	0.304	0.305	-0.3	95	0.00
52 C	Acenaphthene	1.208	1.061	12.2	83	0.00
53	3-Nitroaniline	0.320	0.307	4.1	88	0.00
54 P	2,4-Dinitrophenol	0.176	0.154	12.5	83	0.00
55	Dibenzofuran	1.726	1.681	2.6	93	0.00
56 P	4-Nitrophenol	0.253	0.212	16.2	78	0.00
57	2,4-Dinitrotoluene	0.413	0.400	3.1	90	0.00
58	Fluorene	1.324	1.287	2.8	92	0.00
59	2,3,4,6-Tetrachlorophenol	0.321	0.321	0.0	93	0.00
60	Diethylphthalate	1.432	1.391	2.9	93	0.00
61	4-Chlorophenyl-phenylether	0.640	0.616	3.8	95	0.00
62	4-Nitroaniline	0.317	0.294	7.3	87	0.00
63	Azobenzene	1.267	1.358	-7.2	103	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	0.130	0.116	10.8	76	0.00
66 c	n-Nitrosodiphenylamine	0.648	0.644	0.6	92	0.00
67	4-Bromophenyl-phenylether	0.212	0.214	-0.9	95	0.00
68	Hexachlorobenzene	0.220	0.235	-6.8	100	0.00
69	Atrazine	0.197	0.205	-4.1	97	0.00
70 C	Pentachlorophenol	0.137	0.119	13.1	77	0.00
71	Phenanthrene	1.076	1.081	-0.5	92	0.00
72	Anthracene	1.075	1.079	-0.4	93	0.00
73	Carbazole	1.008	0.966	4.2	89	0.00
74	Di-n-butylphthalate	1.343	1.288	4.1	87	0.00
75 C	Fluoranthene	1.098	1.070	2.6	89	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
77	Benzydine	0.564	0.573	-1.6	83	0.00
78	Pyrene	1.583	1.561	1.4	88	0.00
79 S	Terphenyl-d14	1.167	1.182	-1.3	91	0.00
80	Butylbenzylphthalate	0.758	0.718	5.3	84	0.00
81	Benzo(a)anthracene	1.307	1.245	4.7	86	0.00
82	3,3'-Dichlorobenzidine	0.363	0.334	8.0	82	0.00
83	Chrysene	1.260	1.191	5.5	83	0.00
84	Bis(2-ethylhexyl)phthalate	1.117	1.036	7.3	82	0.00
85 c	Di-n-octyl phthalate	1.813	1.565	13.7	76	0.00
86 I	Perylene-d12	1.000	1.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	1.155	1.160	-0.4	87	0.00
88	Benzo(b)fluoranthene	1.408	1.325	5.9	81	0.00
89	Benzo(k)fluoranthene	1.287	1.262	1.9	83	0.00
90 C	Benzo(a)pyrene	1.176	1.143	2.8	83	0.00
91	Dibenzo(a,h)anthracene	1.011	0.991	2.0	86	0.00
92	Benzo(g,h,i)perylene	0.959	0.961	-0.2	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080221\
Data File : BF124898.D
Acq On : 2 Aug 2021 9:31
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 02 11:23:06 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 11:15:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0