

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080222\
 Data File : BF129514.D
 Acq On : 02 Aug 2022 18:52
 Operator : CG\JU
 Sample : N3927-20
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RVE-93

Quant Time: Aug 03 01:47:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 01:11:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	168674	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	682527	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	355865	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	573806	20.000	ng	0.00
76) Chrysene-d12	14.045	240	313913	20.000	ng	0.00
86) Perylene-d12	15.533	264	324300	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1218437	117.673	ng	0.01
7) Phenol-d6	6.522	99	1552850	119.313	ng	0.00
23) Nitrobenzene-d5	7.439	82	972745	77.800	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	358644	104.824	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	1516852	65.938	ng	0.00
79) Terphenyl-d14	12.986	244	1148059	68.997	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080222\
Data File : BF129514.D
Acq On : 02 Aug 2022 18:52
Operator : CG\JU
Sample : N3927-20
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RVE-93

Quant Time: Aug 03 01:47:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 02 01:11:25 2022
Response via : Initial Calibration

