

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080222\
 Data File : BF129498.D
 Acq On : 02 Aug 2022 09:46
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 17:34:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 01:11:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	76	0.00
2	1,4-Dioxane	40.000	38.625	3.4	74	0.00
3	Pyridine	40.000	38.172	4.6	71	0.00
4	n-Nitrosodimethylamine	40.000	40.183	-0.5	76	0.01
5 S	2-Fluorophenol	80.000	78.923	1.3	77	0.01
6	Aniline	40.000	37.155	7.1	71	0.00
7 S	Phenol-d6	80.000	79.925	0.1	78	0.00
8	2-Chlorophenol	40.000	40.365	-0.9	78	0.00
9	Benzaldehyde	40.000	36.221	9.4	74	0.00
10 C	Phenol	40.000	40.867	-2.2	80	0.01
11	bis(2-Chloroethyl)ether	40.000	39.299	1.8	76	0.00
12	1,3-Dichlorobenzene	40.000	40.679	-1.7	79	0.00
13 C	1,4-Dichlorobenzene	40.000	40.195	-0.5	78	0.00
14	1,2-Dichlorobenzene	40.000	40.652	-1.6	78	0.00
15	Benzyl Alcohol	40.000	42.069	-5.2	80	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.429	8.9	70	0.00
17	2-Methylphenol	40.000	39.806	0.5	78	0.00
18	Hexachloroethane	40.000	40.854	-2.1	79	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.157	-0.4	79	0.00
20	3+4-Methylphenols	40.000	39.901	0.2	81	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	76	0.00
22	Acetophenone	40.000	41.144	-2.9	82	0.00
23 S	Nitrobenzene-d5	80.000	83.178	-4.0	80	0.00
24	Nitrobenzene	40.000	40.705	-1.8	77	0.00
25	Isophorone	40.000	40.126	-0.3	78	0.00
26 C	2-Nitrophenol	40.000	42.485	-6.2	80	0.00
27	2,4-Dimethylphenol	40.000	40.721	-1.8	79	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.146	-0.4	78	0.00
29 C	2,4-Dichlorophenol	40.000	41.001	-2.5	78	0.00
30	1,2,4-Trichlorobenzene	40.000	40.977	-2.4	79	0.00
31	Naphthalene	40.000	40.906	-2.3	80	0.00
32	Benzoic acid	40.000	27.434	31.4#	50	0.01
33	4-Chloroaniline	40.000	40.526	-1.3	79	0.00
34 C	Hexachlorobutadiene	40.000	42.633	-6.6	83	0.00
35	Caprolactam	40.000	42.163	-5.4	81	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.758	-4.4	81	0.00
37	2-Methylnaphthalene	40.000	41.285	-3.2	81	0.00
38	1-Methylnaphthalene	40.000	41.145	-2.9	81	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	77	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.351	-3.4	82	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.898	-2.2	73	0.00
42 S	2,4,6-Tribromophenol	80.000	84.931	-6.2	84	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.711	-1.8	79	0.00
44	2,4,5-Trichlorophenol	40.000	40.345	-0.9	79	0.01
45 S	2-Fluorobiphenyl	80.000	76.562	4.3	84	0.00
46	1,1'-Biphenyl	40.000	40.989	-2.5	81	0.00
47	2-Chloronaphthalene	40.000	40.404	-1.0	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080222\
 Data File : BF129498.D
 Acq On : 02 Aug 2022 09:46
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 17:34:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 01:11:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.959	-4.9	80	0.00
49	Acenaphthylene	40.000	40.644	-1.6	81	0.00
50	Dimethylphthalate	40.000	41.442	-3.6	82	0.00
51	2,6-Dinitrotoluene	40.000	42.041	-5.1	82	0.00
52 C	Acenaphthene	40.000	41.102	-2.8	80	0.00
53	3-Nitroaniline	40.000	40.676	-1.7	78	0.00
54 P	2,4-Dinitrophenol	40.000	42.426	-6.1	78	0.01
55	Dibenzofuran	40.000	40.960	-2.4	81	0.00
56 P	4-Nitrophenol	40.000	38.066	4.8	73	0.02
57	2,4-Dinitrotoluene	40.000	41.900	-4.7	79	0.00
58	Fluorene	40.000	41.821	-4.6	85	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.144	-0.4	77	0.00
60	Diethylphthalate	40.000	41.467	-3.7	82	0.00
61	4-Chlorophenyl-phenylether	40.000	41.625	-4.1	83	0.00
62	4-Nitroaniline	40.000	41.073	-2.7	80	0.00
63	Azobenzene	40.000	40.152	-0.4	79	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	79	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.281	-3.2	77	0.01
66 c	n-Nitrosodiphenylamine	40.000	40.301	-0.8	81	0.01
67	4-Bromophenyl-phenylether	40.000	41.046	-2.6	82	0.00
68	Hexachlorobenzene	40.000	41.526	-3.8	83	0.00
69	Atrazine	40.000	40.002	-0.0	80	0.00
70 C	Pentachlorophenol	40.000	40.409	-1.0	76	0.01
71	Phenanthrene	40.000	39.942	0.1	81	0.00
72	Anthracene	40.000	40.584	-1.5	83	0.01
73	Carbazole	40.000	40.007	-0.0	81	0.01
74	Di-n-butylphthalate	40.000	40.455	-1.1	82	0.00
75 C	Fluoranthene	40.000	40.820	-2.1	83	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	82	0.01
77	Benzidine	40.000	44.246	-10.6	86	0.01
78	Pyrene	40.000	40.415	-1.0	82	0.01
79 S	Terphenyl-d14	80.000	81.947	-2.4	86	0.01
80	Butylbenzylphthalate	40.000	41.505	-3.8	83	0.00
81	Benzo(a)anthracene	40.000	40.521	-1.3	84	0.01
82	3,3'-Dichlorobenzidine	40.000	38.124	4.7	78	0.01
83	Chrysene	40.000	40.206	-0.5	84	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	38.938	2.7	80	0.00
85 c	Di-n-octyl phthalate	40.000	40.423	-1.1	84	0.01
86 I	Perylene-d12	20.000	20.000	0.0	80	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	40.468	-1.2	75	0.03
88	Benzo(b)fluoranthene	40.000	41.238	-3.1	85	0.02
89	Benzo(k)fluoranthene	40.000	40.893	-2.2	81	0.02
90 C	Benzo(a)pyrene	40.000	40.915	-2.3	82	0.02
91	Dibenzo(a,h)anthracene	40.000	40.109	-0.3	73	0.03
92	Benzo(g,h,i)perylene	40.000	40.925	-2.3	74	0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080222\
Data File : BF129498.D
Acq On : 02 Aug 2022 09:46
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 02 17:34:03 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 02 01:11:25 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0