

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080222\
 Data File : BF129510.D
 Acq On : 02 Aug 2022 16:35
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 17:39:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 01:11:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	51	0.00
2	1,4-Dioxane	0.554	0.540	2.5	50#	-0.05
3	Pyridine	1.538	1.519	1.2	49#	-0.04
4	n-Nitrosodimethylamine	0.676	0.668	1.2	50	-0.05
5 S	2-Fluorophenol	1.228	1.271	-3.5	54	0.00
6	Aniline	1.918	1.777	7.4	48#	0.00
7 S	Phenol-d6	1.543	1.582	-2.5	54	0.00
8	2-Chlorophenol	1.341	1.397	-4.2	54	0.00
9	Benzaldehyde	1.048	0.975	7.0	51	0.00
10 C	Phenol	1.643	1.734	-5.5	55	0.00
11	bis(2-Chloroethyl)ether	1.303	1.306	-0.2	52	0.00
12	1,3-Dichlorobenzene	1.439	1.471	-2.2	53	0.00
13 C	1,4-Dichlorobenzene	1.463	1.507	-3.0	54	0.00
14	1,2-Dichlorobenzene	1.345	1.392	-3.5	53	0.00
15	Benzyl Alcohol	1.119	1.197	-7.0	54	0.00
16	2,2'-oxybis(1-Chloropropane	1.959	1.842	6.0	49#	0.00
17	2-Methylphenol	1.113	1.118	-0.4	53	0.00
18	Hexachloroethane	0.551	0.563	-2.2	53	0.00
19 P	n-Nitroso-di-n-propylamine	0.934	0.955	-2.2	54	0.00
20	3+4-Methylphenols	1.415	1.464	-3.5	56	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	51	0.00
22	Acetophenone	0.466	0.490	-5.2	56	0.00
23 S	Nitrobenzene-d5	0.366	0.386	-5.5	55	0.00
24	Nitrobenzene	0.377	0.391	-3.7	53	0.00
25	Isophorone	0.657	0.646	1.7	52	0.00
26 C	2-Nitrophenol	0.186	0.194	-4.3	53	0.00
27	2,4-Dimethylphenol	0.279	0.284	-1.8	53	0.00
28	bis(2-Chloroethoxy)methane	0.381	0.382	-0.3	53	0.00
29 C	2,4-Dichlorophenol	0.288	0.296	-2.8	52	0.00
30	1,2,4-Trichlorobenzene	0.298	0.307	-3.0	54	0.00
31	Naphthalene	1.016	1.059	-4.2	55	0.00
32	Benzoic acid	0.202	0.137	32.2#	33#	0.00
33	4-Chloroaniline	0.417	0.412	1.2	52	0.00
34 C	Hexachlorobutadiene	0.170	0.176	-3.5	54	0.00
35	Caprolactam	0.094	0.100	-6.4	55	-0.01
36 C	4-Chloro-3-methylphenol	0.306	0.318	-3.9	54	0.00
37	2-Methylnaphthalene	0.660	0.688	-4.2	55	0.00
38	1-Methylnaphthalene	0.637	0.662	-3.9	55	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	52	0.00
40	1,2,4,5-Tetrachlorobenzene	0.525	0.543	-3.4	56	0.00
41 P	Hexachlorocyclopentadiene	0.234	0.215	8.1	45#	0.00
42 S	2,4,6-Tribromophenol	0.192	0.205	-6.8	57	0.00
43 C	2,4,6-Trichlorophenol	0.381	0.389	-2.1	54	0.00
44	2,4,5-Trichlorophenol	0.415	0.422	-1.7	54	0.00
45 S	2-Fluorobiphenyl	1.293	1.269	1.9	58	0.00
46	1,1'-Biphenyl	1.460	1.506	-3.2	55	0.00
47	2-Chloronaphthalene	1.180	1.203	-1.9	55	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080222\
 Data File : BF129510.D
 Acq On : 02 Aug 2022 16:35
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 17:39:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 01:11:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.392	0.402	-2.6	53	0.00
49	Acenaphthylene	1.793	1.858	-3.6	56	0.00
50	Dimethylphthalate	1.381	1.418	-2.7	55	0.00
51	2,6-Dinitrotoluene	0.309	0.316	-2.3	54	0.00
52 C	Acenaphthene	1.171	1.193	-1.9	54	0.00
53	3-Nitroaniline	0.365	0.363	0.5	52	0.00
54 P	2,4-Dinitrophenol	0.158	0.140	11.4	44#	0.00
55	Dibenzofuran	1.614	1.676	-3.8	56	0.00
56 P	4-Nitrophenol	0.269	0.260	3.3	50	0.00
57	2,4-Dinitrotoluene	0.404	0.425	-5.2	54	0.00
58	Fluorene	1.292	1.363	-5.5	58	0.00
59	2,3,4,6-Tetrachlorophenol	0.321	0.322	-0.3	52	0.00
60	Diethylphthalate	1.335	1.352	-1.3	54	0.00
61	4-Chlorophenyl-phenylether	0.569	0.589	-3.5	56	0.00
62	4-Nitroaniline	0.362	0.374	-3.3	54	0.00
63	Azobenzene	1.361	1.355	0.4	53	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	53	0.00
65	4,6-Dinitro-2-methylphenol	0.126	0.125	0.8	50#	0.00
66 c	n-Nitrosodiphenylamine	0.666	0.672	-0.9	55	0.00
67	4-Bromophenyl-phenylether	0.219	0.221	-0.9	54	0.00
68	Hexachlorobenzene	0.237	0.247	-4.2	56	0.00
69	Atrazine	0.202	0.200	1.0	53	0.00
70 C	Pentachlorophenol	0.129	0.128	0.8	50	0.00
71	Phenanthrene	1.092	1.123	-2.8	56	0.00
72	Anthracene	1.073	1.120	-4.4	58	0.00
73	Carbazole	1.010	1.058	-4.8	57	0.00
74	Di-n-butylphthalate	1.203	1.247	-3.7	57	0.00
75 C	Fluoranthene	1.106	1.189	-7.5	59	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	61	0.00
77	Benzydine	0.707	0.740	-4.7	61	0.00
78	Pyrene	1.672	1.648	1.4	60	0.00
79 S	Terphenyl-d14	1.060	1.057	0.3	62	0.00
80	Butylbenzylphthalate	0.725	0.724	0.1	59	0.00
81	Benzo(a)anthracene	1.366	1.404	-2.8	63	0.00
82	3,3'-Dichlorobenzidine	0.451	0.456	-1.1	61	0.00
83	Chrysene	1.288	1.323	-2.7	64	0.00
84	Bis(2-ethylhexyl)phthalate	0.955	0.961	-0.6	61	0.00
85 c	Di-n-octyl phthalate	1.374	1.464	-6.6	66	0.00
86 I	Perylene-d12	1.000	1.000	0.0	64	0.00
87	Indeno(1,2,3-cd)pyrene	1.347	1.336	0.8	59	0.01
88	Benzo(b)fluoranthene	1.327	1.309	1.4	66	0.00
89	Benzo(k)fluoranthene	1.300	1.342	-3.2	66	0.00
90 C	Benzo(a)pyrene	1.077	1.089	-1.1	65	0.00
91	Dibenzo(a,h)anthracene	1.096	1.092	0.4	59	0.01
92	Benzo(g,h,i)perylene	1.120	1.112	0.7	58	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080222\
Data File : BF129510.D
Acq On : 02 Aug 2022 16:35
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 02 17:39:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 02 01:11:25 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0