

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080224\
 Data File : BF138736.D
 Acq On : 02 Aug 2024 09:13
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 09:38:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:50:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	114	0.00
2	1,4-Dioxane	40.000	39.107	2.2	110	0.04
3	Pyridine	40.000	39.805	0.5	113	0.04
4	n-Nitrosodimethylamine	40.000	38.891	2.8	111	0.04
5 S	2-Fluorophenol	80.000	78.535	1.8	114	0.00
6	Aniline	40.000	39.514	1.2	114	0.00
7 S	Phenol-d6	80.000	79.153	1.1	116	0.00
8	2-Chlorophenol	40.000	39.541	1.1	116	0.00
9	Benzaldehyde	40.000	46.091	-15.2	151	0.00
10 C	Phenol	40.000	39.141	2.1	115	0.00
11	bis(2-Chloroethyl)ether	40.000	38.701	3.2	114	0.00
12	1,3-Dichlorobenzene	40.000	39.056	2.4	115	0.00
13 C	1,4-Dichlorobenzene	40.000	38.962	2.6	114	0.00
14	1,2-Dichlorobenzene	40.000	38.660	3.4	112	0.00
15	Benzyl Alcohol	40.000	40.569	-1.4	119	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.874	7.8	108	0.00
17	2-Methylphenol	40.000	39.336	1.7	114	0.00
18	Hexachloroethane	40.000	38.067	4.8	111	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.052	2.4	118	0.00
20	3+4-Methylphenols	40.000	39.010	2.5	117	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	116	0.00
22	Acetophenone	40.000	39.677	0.8	117	0.00
23 S	Nitrobenzene-d5	80.000	79.324	0.8	116	0.00
24	Nitrobenzene	40.000	39.371	1.6	114	0.00
25	Isophorone	40.000	39.015	2.5	117	0.00
26 C	2-Nitrophenol	40.000	41.062	-2.7	117	0.00
27	2,4-Dimethylphenol	40.000	39.840	0.4	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.903	2.7	115	0.00
29 C	2,4-Dichlorophenol	40.000	40.648	-1.6	118	0.00
30	1,2,4-Trichlorobenzene	40.000	39.097	2.3	114	0.00
31	Naphthalene	40.000	38.965	2.6	114	0.00
32	Benzoic acid	40.000	40.251	-0.6	119	0.02
33	4-Chloroaniline	40.000	39.897	0.3	119	0.00
34 C	Hexachlorobutadiene	40.000	39.519	1.2	117	0.00
35	Caprolactam	40.000	44.049	-10.1	135	0.01
36 C	4-Chloro-3-methylphenol	40.000	41.150	-2.9	123	0.00
37	2-Methylnaphthalene	40.000	39.849	0.4	119	0.00
38	1-Methylnaphthalene	40.000	39.419	1.5	118	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.928	2.7	119	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.192	7.0	116	0.00
42 S	2,4,6-Tribromophenol	80.000	84.867	-6.1	132	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.988	0.0	122	0.00
44	2,4,5-Trichlorophenol	40.000	39.584	1.0	120	0.00
45 S	2-Fluorobiphenyl	80.000	77.195	3.5	119	0.00
46	1,1'-Biphenyl	40.000	39.387	1.5	121	0.00
47	2-Chloronaphthalene	40.000	39.490	1.3	121	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080224\
 Data File : BF138736.D
 Acq On : 02 Aug 2024 09:13
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 09:38:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:50:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.392	-3.5	127	0.00
49	Acenaphthylene	40.000	39.533	1.2	121	0.00
50	Dimethylphthalate	40.000	40.547	-1.4	127	0.00
51	2,6-Dinitrotoluene	40.000	43.154	-7.9	131	0.00
52 C	Acenaphthene	40.000	39.095	2.3	121	0.00
53	3-Nitroaniline	40.000	41.617	-4.0	130	0.00
54 P	2,4-Dinitrophenol	40.000	43.368	-8.4	137	0.00
55	Dibenzofuran	40.000	39.519	1.2	124	0.00
56 P	4-Nitrophenol	40.000	41.890	-4.7	130	0.00
57	2,4-Dinitrotoluene	40.000	43.832	-9.6	135	0.00
58	Fluorene	40.000	40.310	-0.8	126	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.461	-1.2	126	0.00
60	Diethylphthalate	40.000	40.987	-2.5	130	0.00
61	4-Chlorophenyl-phenylether	40.000	39.661	0.8	125	0.00
62	4-Nitroaniline	40.000	44.895	-12.2	141	0.00
63	Azobenzene	40.000	40.042	-0.1	125	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	135	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.716	-4.3	138	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.145	2.1	131	0.00
67	4-Bromophenyl-phenylether	40.000	38.304	4.2	129	0.00
68	Hexachlorobenzene	40.000	39.113	2.2	134	0.00
69	Atrazine	40.000	39.658	0.9	133	0.00
70 C	Pentachlorophenol	40.000	38.868	2.8	129	0.00
71	Phenanthrene	40.000	39.267	1.8	134	0.00
72	Anthracene	40.000	39.684	0.8	135	0.00
73	Carbazole	40.000	41.097	-2.7	139	0.00
74	Di-n-butylphthalate	40.000	43.598	-9.0	146	0.00
75 C	Fluoranthene	40.000	42.039	-5.1	141	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
77	Benzydine	40.000	37.029	7.4	92	0.00
78	Pyrene	40.000	46.834	-17.1	140	0.00
79 S	Terphenyl-d14	80.000	92.434	-15.5	139	0.00
80	Butylbenzylphthalate	40.000	41.768	-4.4	112	0.00
81	Benzo(a)anthracene	40.000	41.662	-4.2	114	0.00
82	3,3'-Dichlorobenzidine	40.000	39.934	0.2	109	0.00
83	Chrysene	40.000	40.241	-0.6	114	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	33.898	15.3	88	0.00
85 c	Di-n-octyl phthalate	40.000	34.078	14.8	89	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.936	2.7	96	0.00
88	Benzo(b)fluoranthene	40.000	38.884	2.8	94	0.00
89	Benzo(k)fluoranthene	40.000	40.209	-0.5	103	0.00
90 C	Benzo(a)pyrene	40.000	39.791	0.5	97	0.00
91	Dibenzo(a,h)anthracene	40.000	37.876	5.3	94	0.00
92	Benzo(g,h,i)perylene	40.000	37.958	5.1	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080224\
Data File : BF138736.D
Acq On : 02 Aug 2024 09:13
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 02 09:38:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0