

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080318\
 Data File : BF107948.D
 Acq On : 3 Aug 2018 20:00
 Operator : JU/SJ
 Sample : J4318-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BEEBEE-SITE-DEMO-2A

Manual Integrations
 APPROVED

Sohil
 8/6/2018 5:05:40 PM

Quant Time: Aug 06 11:05:18 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 17:48:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	117890	20.00	ng	-0.01
21) Naphthalene-d8	8.24	136	524390	20.00	ng	-0.01
38) Acenaphthene-d10	9.99	164	214564	20.00	ng	-0.02
63) Phenanthrene-d10	11.49	188	368381	20.00	ng	-0.01
75) Chrysene-d12	14.14	240	335172	20.00	ng	-0.01
86) Perylene-d12	15.67	264	367532	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	626696	90.35	ng	0.00
7) Phenol-d6	6.60	99	911926	99.75	ng	0.00
23) Nitrobenzene-d5	7.52	82	510618	56.10	ng	-0.02
41) 2,4,6-Tribromophenol	10.79	330	221136	75.85	ng	-0.01
44) 2-Fluorobiphenyl	9.30	172	1111536	72.19	ng	-0.02
78) Terphenyl-d14	13.07	244	888434	57.48	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.70	163	68844	4.272	ng	# 98
74) Fluoranthene	12.71	202	79756	3.881	ng	97
77) Pyrene	12.94	202	73524	3.046	ng	97
87) Benzo(b)fluoranthene	15.22	252	56684	3.127	ng	# 87
88) Benzo(k)fluoranthene	15.22	252	56684	2.547	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080318\
Data File : BF107948.D
Acq On : 3 Aug 2018 20:00
Operator : JU/SJ
Sample : J4318-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BEEBEE-SITE-DEMO-2A

Manual Integrations
APPROVED

Sohil
8/6/2018 5:05:40 PM

Quant Time: Aug 06 11:05:18 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 30 17:48:53 2018
Response via : Initial Calibration

