


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080318\
 Data File : BF107960.D
 Acq On : 4 Aug 2018 1:52
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF073018.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jul 30 17:48:53 2018



AutoFind: Scans 1593, 1594, 1595; Background Corrected with Scan 1588

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.2	22319	PASS
68	69	0.00	2	1.6	295	PASS
69	198	0.00	100	33.7	18287	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.8	24817	PASS
197	198	0.00	2	0.4	203	PASS
198	198	100	100	100.0	54216	PASS
199	198	5	9	7.1	3862	PASS
275	198	10	60	27.0	14634	PASS
365	198	1	100	3.5	1918	PASS
441	198	0.01	100	14.7	7964	PASS
442	198	50	100	92.6	50221	PASS
443	442	15	24	19.4	9728	PASS